

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022423\
 Data File : P0092726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2023 22:04
 Operator : YP/AJ
 Sample : PB151063BL
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 04:23:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.431	3.637	64580848	25439117	21.507	19.957
2)	SA Decachlor...	10.262	8.670	46366648	24229378	24.106	20.791
Target Compounds							
3)	L1 AR-1016-1	5.542f	0.000	1123233	0	12.221	N.D. #
9)	L2 AR-1221-2	4.738	3.935	879441	254403	30.565	21.013 #
10)	L2 AR-1221-3	4.738f	0.000	879441	0	9.986	N.D. #
11)	L3 AR-1232-1	4.738f	0.000	879441	0	12.515	N.D. #
16)	L4 AR-1242-1	5.542f	0.000	1123233	0	14.236	N.D. #
21)	L5 AR-1248-1	5.542f	0.000	1123233	0	18.323	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022423\
Data File : PO092726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2023 22:04
Operator : YP/AJ
Sample : PB151063BL
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 04:23:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: -0.003 min
Response: 64580848
Conc: 21.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.001 min
Response: 25439117
Conc: 19.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.262 min
Delta R.T.: 0.006 min
Response: 46366648
Conc: 24.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.670 min
Delta R.T.: -0.002 min
Response: 24229378
Conc: 20.79 ng/ml



#3 AR-1016-1


R.T.: 5.542 min
Delta R.T.: -0.067 min
Response: 1123233
Conc: 12.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.725 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.738 min
Delta R.T.: 0.009 min
Response: 879441
Conc: 30.56 ng/ml



#9 AR-1221-2


R.T.: 3.935 min
Delta R.T.: -0.002 min
Response: 254403
Conc: 21.01 ng/ml



#10 AR-1221-3


R.T.: 4.738 min
Delta R.T.: -0.067 min
Response: 879441
Conc: 9.99 ng/ml

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.014 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 4.738 min
Delta R.T.: -0.066 min
Response: 879441
Conc: 12.51 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.012 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.542 min
Delta R.T.: -0.068 min
Response: 1123233
Conc: 14.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 4.724 min
Response: 0
Conc: N.D.



#21 AR-1248-1

R.T.: 5.542 min
Delta R.T.: -0.068 min
Response: 1123233
Conc: 18.32 ng/ml



#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.724 min
Response: 0
Conc: N.D.