

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022624\
 Data File : P0102092.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2024 20:19
 Operator : YP/AJ
 Sample : P1564-13
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LKWD-BACKFILL-15

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/27/2024
 Supervised By :Ankita Jodhani 02/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 00:13:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.278	3.390	97294276	66357296	20.980	20.421
2) SA Decachlor...	9.830	8.266	50632087	39469591	19.884m	19.079

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022624\
Data File : PO102092.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2024 20:19
Operator : YP/AJ
Sample : P1564-13
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
LKWD-BACKFILL-15


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/27/2024
Supervised By :Ankita Jodhani 02/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 00:13:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.278 min
Delta R.T.: -0.004 min
Response: 97294276
Conc: 20.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
LKWD-BACKFILL-15


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/27/2024
Supervised By :Ankita Jodhani 02/27/2024




#1 Tetrachloro-m-xylene

R.T.: 3.390 min
Delta R.T.: -0.002 min
Response: 66357296
Conc: 20.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.830 min
Delta R.T.: -0.011 min
Response: 50632087
Conc: 19.88 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.266 min
Delta R.T.: -0.005 min
Response: 39469591
Conc: 19.08 ng/ml