

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022723\
 Data File : P0092792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2023 18:12
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 20:49:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.639	161.6E6	62571920	53.832	49.088
2) SA Decachlor...	10.274	8.677	102.5E6	52234711	53.304	44.821
Target Compounds						
16) L4 AR-1242-1	5.610	4.726	40258512	16584078	510.239	493.851
17) L4 AR-1242-2	5.633	4.744	59154527	23693011	514.245	494.092
18) L4 AR-1242-3	5.696	4.921	37924593	13080011	513.724	502.309
19) L4 AR-1242-4	5.795	5.005	28842274	14331780	520.743	497.019
20) L4 AR-1242-5	6.539	5.530	29864216	15848772	509.864	506.527



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022723\
Data File : PO092792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2023 18:12
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 20:49:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.433 min
Delta R.T.: 0.000 min
Response: 161644836
Conc: 53.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 62571920
Conc: 49.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: 0.017 min
Response: 102529908
Conc: 53.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.005 min
Response: 52234711
Conc: 44.82 ng/ml



#16 AR-1242-1


R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 40258512
Conc: 510.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.726 min
Delta R.T.: 0.002 min
Response: 16584078
Conc: 493.85 ng/ml




#17 AR-1242-2

R.T.: 5.633 min
Delta R.T.: 0.000 min
Response: 59154527
Conc: 514.24 ng/ml



#17 AR-1242-2


R.T.: 4.744 min
Delta R.T.: 0.002 min
Response: 23693011
Conc: 494.09 ng/ml



#18 AR-1242-3


R.T.: 5.696 min
Delta R.T.: 0.002 min
Response: 37924593
Conc: 513.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: 0.002 min
Response: 13080011
Conc: 502.31 ng/ml




#19 AR-1242-4

R.T.: 5.795 min
Delta R.T.: 0.002 min
Response: 28842274
Conc: 520.74 ng/ml



#19 AR-1242-4


R.T.: 5.005 min
Delta R.T.: 0.002 min
Response: 14331780
Conc: 497.02 ng/ml



#20 AR-1242-5

R.T.: 6.539 min
Delta R.T.: 0.003 min
Response: 29864216
Conc: 509.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.530 min
Delta R.T.: 0.003 min
Response: 15848772
Conc: 506.53 ng/ml