

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022723\
 Data File : P0092793.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2023 18:29
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 20:50:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.639	157.8E6	61417648	52.559	48.183
2) SA Decachlor...	10.273	8.677	101.2E6	52463700	52.603	45.018
Target Compounds						
21) L5 AR-1248-1	5.611	4.726	30492005	12868217	497.418	491.442
22) L5 AR-1248-2	5.887	4.964	47196790	19336536	490.415	477.934
23) L5 AR-1248-3	6.093	5.006	50408976	20546454	491.229	487.510
24) L5 AR-1248-4	6.500	5.177	48393100	23520473	483.550	486.936
25) L5 AR-1248-5	6.539	5.572	49346404	20834139	487.474	499.179



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022723\
Data File : P0092793.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2023 18:29
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 20:50:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 157820542
Conc: 52.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 61417648
Conc: 48.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.273 min
Delta R.T.: 0.017 min
Response: 101181239
Conc: 52.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.005 min
Response: 52463700
Conc: 45.02 ng/ml



#21 AR-1248-1


R.T.: 5.611 min
Delta R.T.: 0.001 min
Response: 30492005
Conc: 497.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.726 min
Delta R.T.: 0.002 min
Response: 12868217
Conc: 491.44 ng/ml




#22 AR-1248-2

R.T.: 5.887 min
Delta R.T.: 0.001 min
Response: 47196790
Conc: 490.41 ng/ml



#22 AR-1248-2


R.T.: 4.964 min
Delta R.T.: 0.002 min
Response: 19336536
Conc: 477.93 ng/ml



#23 AR-1248-3


R.T.: 6.093 min
Delta R.T.: 0.001 min
Response: 50408976
Conc: 491.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.006 min
Delta R.T.: 0.002 min
Response: 20546454
Conc: 487.51 ng/ml




#24 AR-1248-4

R.T.: 6.500 min
Delta R.T.: 0.001 min
Response: 48393100
Conc: 483.55 ng/ml



#24 AR-1248-4


R.T.: 5.177 min
Delta R.T.: 0.002 min
Response: 23520473
Conc: 486.94 ng/ml



#25 AR-1248-5

R.T.: 6.539 min
Delta R.T.: 0.002 min
Response: 49346404
Conc: 487.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.572 min
Delta R.T.: 0.002 min
Response: 20834139
Conc: 499.18 ng/ml