

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022724\
 Data File : P0102134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2024 18:33
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 00:27:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.279	3.391	223.0E6	151.5E6	48.086	46.628
2) SA Decachlor...	9.831	8.266	124.3E6	92862804	48.819	44.888
Target Compounds						
21) L5 AR-1248-1	5.424	4.445	36447200	27136041	487.657	492.166
22) L5 AR-1248-2	5.694	4.676	57575057	42087880	496.306	501.450
23) L5 AR-1248-3	5.894	4.716	59748442	42778138	490.204	487.202
24) L5 AR-1248-4	6.293	4.884	61333612	49247803	508.261	489.016
25) L5 AR-1248-5	6.331	5.269	59708208	43604396	500.292	498.707



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022724\
Data File : PO102134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2024 18:33
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 00:27:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 223000680
Conc: 48.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.391 min
Delta R.T.: -0.001 min
Response: 151513876
Conc: 46.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.831 min
Delta R.T.: -0.010 min
Response: 124312620
Conc: 48.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.266 min
Delta R.T.: -0.006 min
Response: 92862804
Conc: 44.89 ng/ml



#21 AR-1248-1


R.T.: 5.424 min
Delta R.T.: -0.003 min
Response: 36447200
Conc: 487.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.445 min
Delta R.T.: -0.002 min
Response: 27136041
Conc: 492.17 ng/ml




#22 AR-1248-2

R.T.: 5.694 min
Delta R.T.: -0.004 min
Response: 57575057
Conc: 496.31 ng/ml



#22 AR-1248-2


R.T.: 4.676 min
Delta R.T.: -0.002 min
Response: 42087880
Conc: 501.45 ng/ml



#23 AR-1248-3


R.T.: 5.894 min
Delta R.T.: -0.004 min
Response: 59748442
Conc: 490.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.716 min
Delta R.T.: -0.002 min
Response: 42778138
Conc: 487.20 ng/ml




#24 AR-1248-4

R.T.: 6.293 min
Delta R.T.: -0.005 min
Response: 61333612
Conc: 508.26 ng/ml



#24 AR-1248-4


R.T.: 4.884 min
Delta R.T.: -0.002 min
Response: 49247803
Conc: 489.02 ng/ml



#25 AR-1248-5

R.T.: 6.331 min
Delta R.T.: -0.004 min
Response: 59708208
Conc: 500.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.269 min
Delta R.T.: -0.003 min
Response: 43604396
Conc: 498.71 ng/ml