

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022724\
 Data File : P0102164.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2024 04:34
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 05:06:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.279	3.391	218.2E6	149.5E6	47.060	46.019
2) SA Decachlor...	9.830	8.265	123.8E6	94100824	48.604	45.486
Target Compounds						
21) L5 AR-1248-1	5.424	4.444	35667105	27851810	477.219	505.148
22) L5 AR-1248-2	5.694	4.676	57605860	42659425	496.571	508.259
23) L5 AR-1248-3	5.895	4.716	59777054	43979618	490.438	500.886
24) L5 AR-1248-4	6.293	4.884	61000881	50002779	505.504	496.513
25) L5 AR-1248-5	6.331	5.269	61109828	43603646	512.036	498.698



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022724\
Data File : PO102164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2024 04:34
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 05:06:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 218243424
Conc: 47.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.391 min
Delta R.T.: 0.000 min
Response: 149537108
Conc: 46.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.830 min
Delta R.T.: -0.011 min
Response: 123765707
Conc: 48.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.265 min
Delta R.T.: -0.007 min
Response: 94100824
Conc: 45.49 ng/ml



#21 AR-1248-1


R.T.: 5.424 min
Delta R.T.: -0.003 min
Response: 35667105
Conc: 477.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.444 min
Delta R.T.: -0.002 min
Response: 27851810
Conc: 505.15 ng/ml




#22 AR-1248-2

R.T.: 5.694 min
Delta R.T.: -0.004 min
Response: 57605860
Conc: 496.57 ng/ml



#22 AR-1248-2


R.T.: 4.676 min
Delta R.T.: -0.002 min
Response: 42659425
Conc: 508.26 ng/ml



#23 AR-1248-3


R.T.: 5.895 min
Delta R.T.: -0.004 min
Response: 59777054
Conc: 490.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.716 min
Delta R.T.: -0.003 min
Response: 43979618
Conc: 500.89 ng/ml




#24 AR-1248-4

R.T.: 6.293 min
Delta R.T.: -0.004 min
Response: 61000881
Conc: 505.50 ng/ml



#24 AR-1248-4


R.T.: 4.884 min
Delta R.T.: -0.003 min
Response: 50002779
Conc: 496.51 ng/ml



#25 AR-1248-5

R.T.: 6.331 min
Delta R.T.: -0.004 min
Response: 61109828
Conc: 512.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.269 min
Delta R.T.: -0.003 min
Response: 43603646
Conc: 498.70 ng/ml