

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022725\
 Data File : P0109517.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2025 10:13
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 15:32:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 04:40:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.668f	3.690	147675	56676	0.016	0.011 #
2)	SA Decachlor...	8.754	8.699	280261	194132	0.033	0.061 #
Target Compounds							
3)	L1 AR-1016-1	4.799	4.789	103964	195371	0.337	1.251 #
4)	L1 AR-1016-2	4.820	4.803	73820	76454	0.175	0.353 #
5)	L1 AR-1016-3	4.863	4.967	80531	146706	0.272	1.228 #
6)	L1 AR-1016-4	4.974	4.997	79099	71936	0.341	0.691 #
7)	L1 AR-1016-5	5.241	5.226	387773	574733	1.502	4.218 #
8)	L2 AR-1221-1	3.915	3.908	134006	205568	1.016	3.001 #
9)	L2 AR-1221-2	4.005	3.993	134719	133308	1.365	2.597 #
10)	L2 AR-1221-3	4.068	4.064	105098	115052	0.389	0.791 #
11)	L3 AR-1232-1	4.068	4.064	105098	115052	0.503	1.032 #
12)	L3 AR-1232-2	4.584	4.803	345045	76454	2.972	0.740 #
13)	L3 AR-1232-3	4.820	4.967	73820	146706	0.366	2.584 #
14)	L3 AR-1232-4	4.974	5.050	79099	213275	0.711	3.885 #
15)	L3 AR-1232-5	5.061	5.226	410897	574733	5.031	9.774 #
16)	L4 AR-1242-1	4.799	4.789	103964	195371	0.397	1.476 #
17)	L4 AR-1242-2	4.820	4.803	73820	76454	0.207	0.421 #
18)	L4 AR-1242-3	4.863	4.967	80531	146706	0.320	1.469 #
19)	L4 AR-1242-4	4.974	5.050	79099	213275	0.403	1.995 #
20)	L4 AR-1242-5	5.637	5.565	2839327	5389481	13.245	42.874 #
21)	L5 AR-1248-1	4.799	4.789	103964	195371	0.521	1.943 #
22)	L5 AR-1248-2	5.061	4.997	410897	71936	1.477	0.494 #
23)	L5 AR-1248-3	5.241	5.050	387773	213275	1.110	1.376
24)	L5 AR-1248-4	5.603	5.226	2914977	574733	6.110	3.179 #
25)	L5 AR-1248-5	5.637	5.624	2839327	1864859	8.290	10.829 #
26)	L6 AR-1254-1	5.603	5.565	2914977	5389481	5.653	20.607 #
27)	L6 AR-1254-2	5.773	5.720	793693	466973	1.762	1.991
28)	L6 AR-1254-3	6.162	6.136	1013105	558120	1.415	1.565
29)	L6 AR-1254-4	6.379	6.336	749347	348862	1.789	1.806
30)	L6 AR-1254-5	6.793	6.777	110461	123081	0.177	0.415 #
31)	L7 AR-1260-1	6.284	6.253	329277	754111	0.706	3.183 #
32)	L7 AR-1260-2	6.500	6.422	272989	880444	0.483	3.200 #
33)	L7 AR-1260-3	6.849	6.607	207999	266932	0.437	1.047 #
34)	L7 AR-1260-4	7.129	7.060	72702	354628	0.168	1.720 #
35)	L7 AR-1260-5	7.333	7.303	118713	419949	0.119	0.938 #
36)	L8 AR-1262-1	6.840	6.807	119573	188571	0.180	0.594 #
37)	L8 AR-1262-2	7.333	7.303	118713	419949	0.105	0.842 #
38)	L8 AR-1262-3	7.640	7.619	154424	364595	0.344	1.858 #
39)	L8 AR-1262-4	7.690	7.659	156071	332291	0.188	0.936 #
40)	L8 AR-1262-5	8.199	8.091f	84641	64874	0.227	0.453 #
41)	L9 AR-1268-1	7.640	7.619	154424	364595	0.117	0.638 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022725\
Data File : P0109517.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2025 10:13
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 15:32:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	7.690	7.659	156071	332291	0.129	0.633 #
43)	L9 AR-1268-3	0.000	7.860	0	72406	N.D.	0.169 #
44)	L9 AR-1268-4	8.199	8.091f	84641	64874	0.199	0.402 #
45)	L9 AR-1268-5	8.500	8.450	116273	55286	0.039	0.051 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022725\
Data File : PO109517.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2025 10:13
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 15:32:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.668 min
Delta R.T.: -0.030 min
Response: 147675
Conc: 0.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: -0.004 min
Response: 56676
Conc: 0.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.754 min
Delta R.T.: -0.003 min
Response: 280261
Conc: 0.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.699 min
Delta R.T.: -0.008 min
Response: 194132
Conc: 0.06 ng/ml



#3 AR-1016-1


R.T.: 4.799 min
Delta R.T.: 0.007 min
Response: 103964
Conc: 0.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#3 AR-1016-1

R.T.: 4.789 min
Delta R.T.: 0.011 min
Response: 195371
Conc: 1.25 ng/ml




#4 AR-1016-2

R.T.: 4.820 min
Delta R.T.: 0.009 min
Response: 73820
Conc: 0.18 ng/ml



#4 AR-1016-2


R.T.: 4.803 min
Delta R.T.: 0.006 min
Response: 76454
Conc: 0.35 ng/ml



#5 AR-1016-3


R.T.: 4.863 min
Delta R.T.: -0.004 min
Response: 80531
Conc: 0.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: -0.006 min
Response: 146706
Conc: 1.23 ng/ml




#6 AR-1016-4

R.T.: 4.974 min
Delta R.T.: -0.014 min
Response: 79099
Conc: 0.34 ng/ml



#6 AR-1016-4


R.T.: 4.997 min
Delta R.T.: -0.017 min
Response: 71936
Conc: 0.69 ng/ml



#7 AR-1016-5


R.T.: 5.241 min
Delta R.T.: -0.005 min
Response: 387773
Conc: 1.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#7 AR-1016-5

R.T.: 5.226 min
Delta R.T.: -0.002 min
Response: 574733
Conc: 4.22 ng/ml




#8 AR-1221-1

R.T.: 3.915 min
Delta R.T.: 0.005 min
Response: 134006
Conc: 1.02 ng/ml



#8 AR-1221-1


R.T.: 3.908 min
Delta R.T.: 0.003 min
Response: 205568
Conc: 3.00 ng/ml



#9 AR-1221-2


R.T.: 4.005 min
Delta R.T.: 0.008 min
Response: 134719
Conc: 1.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#9 AR-1221-2

R.T.: 3.993 min
Delta R.T.: 0.003 min
Response: 133308
Conc: 2.60 ng/ml




#10 AR-1221-3

R.T.: 4.068 min
Delta R.T.: -0.006 min
Response: 105098
Conc: 0.39 ng/ml



#10 AR-1221-3


R.T.: 4.064 min
Delta R.T.: -0.003 min
Response: 115052
Conc: 0.79 ng/ml



#11 AR-1232-1


R.T.: 4.068 min
Delta R.T.: -0.006 min
Response: 105098
Conc: 0.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#11 AR-1232-1

R.T.: 4.064 min
Delta R.T.: -0.003 min
Response: 115052
Conc: 1.03 ng/ml




#12 AR-1232-2

R.T.: 4.584 min
Delta R.T.: 0.016 min
Response: 345045
Conc: 2.97 ng/ml



#12 AR-1232-2


R.T.: 4.803 min
Delta R.T.: 0.007 min
Response: 76454
Conc: 0.74 ng/ml



#13 AR-1232-3


R.T.: 4.820 min
Delta R.T.: 0.009 min
Response: 73820
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#13 AR-1232-3

R.T.: 4.967 min
Delta R.T.: -0.005 min
Response: 146706
Conc: 2.58 ng/ml




#14 AR-1232-4

R.T.: 4.974 min
Delta R.T.: -0.014 min
Response: 79099
Conc: 0.71 ng/ml



#14 AR-1232-4


R.T.: 5.050 min
Delta R.T.: -0.006 min
Response: 213275
Conc: 3.89 ng/ml



#15 AR-1232-5


R.T.: 5.061 min
Delta R.T.: 0.031 min
Response: 410897
Conc: 5.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#15 AR-1232-5

R.T.: 5.226 min
Delta R.T.: -0.001 min
Response: 574733
Conc: 9.77 ng/ml




#16 AR-1242-1

R.T.: 4.799 min
Delta R.T.: 0.007 min
Response: 103964
Conc: 0.40 ng/ml



#16 AR-1242-1


R.T.: 4.789 min
Delta R.T.: 0.012 min
Response: 195371
Conc: 1.48 ng/ml



#17 AR-1242-2


R.T.: 4.820 min
Delta R.T.: 0.009 min
Response: 73820
Conc: 0.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#17 AR-1242-2

R.T.: 4.803 min
Delta R.T.: 0.006 min
Response: 76454
Conc: 0.42 ng/ml




#18 AR-1242-3

R.T.: 4.863 min
Delta R.T.: -0.004 min
Response: 80531
Conc: 0.32 ng/ml



#18 AR-1242-3


R.T.: 4.967 min
Delta R.T.: -0.006 min
Response: 146706
Conc: 1.47 ng/ml



#19 AR-1242-4


R.T.: 4.974 min
Delta R.T.: -0.015 min
Response: 79099
Conc: 0.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#19 AR-1242-4

R.T.: 5.050 min
Delta R.T.: -0.007 min
Response: 213275
Conc: 1.99 ng/ml




#20 AR-1242-5

R.T.: 5.637 min
Delta R.T.: -0.005 min
Response: 2839327
Conc: 13.24 ng/ml



#20 AR-1242-5


R.T.: 5.565 min
Delta R.T.: -0.013 min
Response: 5389481
Conc: 42.87 ng/ml



#21 AR-1248-1


R.T.: 4.799 min
Delta R.T.: 0.007 min
Response: 103964
Conc: 0.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#21 AR-1248-1

R.T.: 4.789 min
Delta R.T.: 0.012 min
Response: 195371
Conc: 1.94 ng/ml




#22 AR-1248-2

R.T.: 5.061 min
Delta R.T.: 0.031 min
Response: 410897
Conc: 1.48 ng/ml



#22 AR-1248-2


R.T.: 4.997 min
Delta R.T.: -0.016 min
Response: 71936
Conc: 0.49 ng/ml



#23 AR-1248-3


R.T.: 5.241 min
Delta R.T.: -0.005 min
Response: 387773
Conc: 1.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: -0.006 min
Response: 213275
Conc: 1.38 ng/ml




#24 AR-1248-4

R.T.: 5.603 min
Delta R.T.: 0.003 min
Response: 2914977
Conc: 6.11 ng/ml



#24 AR-1248-4


R.T.: 5.226 min
Delta R.T.: -0.001 min
Response: 574733
Conc: 3.18 ng/ml



#25 AR-1248-5


R.T.: 5.637 min
Delta R.T.: -0.005 min
Response: 2839327
Conc: 8.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 5.624 min
Delta R.T.: 0.004 min
Response: 1864859
Conc: 10.83 ng/ml




#26 AR-1254-1

R.T.: 5.603 min
Delta R.T.: 0.003 min
Response: 2914977
Conc: 5.65 ng/ml



#26 AR-1254-1


R.T.: 5.565 min
Delta R.T.: -0.013 min
Response: 5389481
Conc: 20.61 ng/ml



#27 AR-1254-2


R.T.: 5.773 min
Delta R.T.: 0.024 min
Response: 793693
Conc: 1.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#27 AR-1254-2

R.T.: 5.720 min
Delta R.T.: -0.006 min
Response: 466973
Conc: 1.99 ng/ml




#28 AR-1254-3

R.T.: 6.162 min
Delta R.T.: 0.007 min
Response: 1013105
Conc: 1.42 ng/ml



#28 AR-1254-3


R.T.: 6.136 min
Delta R.T.: 0.007 min
Response: 558120
Conc: 1.56 ng/ml



#29 AR-1254-4


R.T.: 6.379 min
Delta R.T.: -0.006 min
Response: 749347
Conc: 1.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#29 AR-1254-4

R.T.: 6.336 min
Delta R.T.: -0.020 min
Response: 348862
Conc: 1.81 ng/ml




#30 AR-1254-5

R.T.: 6.793 min
Delta R.T.: -0.012 min
Response: 110461
Conc: 0.18 ng/ml



#30 AR-1254-5


R.T.: 6.777 min
Delta R.T.: 0.003 min
Response: 123081
Conc: 0.41 ng/ml



#31 AR-1260-1


R.T.: 6.284 min
Delta R.T.: -0.004 min
Response: 329277
Conc: 0.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#31 AR-1260-1

R.T.: 6.253 min
Delta R.T.: -0.008 min
Response: 754111
Conc: 3.18 ng/ml




#32 AR-1260-2

R.T.: 6.500 min
Delta R.T.: 0.023 min
Response: 272989
Conc: 0.48 ng/ml



#32 AR-1260-2


R.T.: 6.422 min
Delta R.T.: -0.027 min
Response: 880444
Conc: 3.20 ng/ml



#33 AR-1260-3


R.T.: 6.849 min
Delta R.T.: 0.004 min
Response: 207999
Conc: 0.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#33 AR-1260-3

R.T.: 6.607 min
Delta R.T.: 0.005 min
Response: 266932
Conc: 1.05 ng/ml




#34 AR-1260-4

R.T.: 7.129 min
Delta R.T.: 0.023 min
Response: 72702
Conc: 0.17 ng/ml



#34 AR-1260-4


R.T.: 7.060 min
Delta R.T.: -0.014 min
Response: 354628
Conc: 1.72 ng/ml



#35 AR-1260-5


R.T.: 7.333 min
Delta R.T.: -0.014 min
Response: 118713
Conc: 0.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: -0.011 min
Response: 419949
Conc: 0.94 ng/ml




#36 AR-1262-1

R.T.: 6.840 min
Delta R.T.: -0.005 min
Response: 119573
Conc: 0.18 ng/ml



#36 AR-1262-1


R.T.: 6.807 min
Delta R.T.: -0.008 min
Response: 188571
Conc: 0.59 ng/ml



#37 AR-1262-2


R.T.: 7.333 min
Delta R.T.: -0.013 min
Response: 118713
Conc: 0.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#37 AR-1262-2

R.T.: 7.303 min
Delta R.T.: -0.010 min
Response: 419949
Conc: 0.84 ng/ml




#38 AR-1262-3

R.T.: 7.640 min
Delta R.T.: 0.008 min
Response: 154424
Conc: 0.34 ng/ml



#38 AR-1262-3


R.T.: 7.619 min
Delta R.T.: 0.022 min
Response: 364595
Conc: 1.86 ng/ml



#39 AR-1262-4


R.T.: 7.690 min
Delta R.T.: -0.006 min
Response: 156071
Conc: 0.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#39 AR-1262-4

R.T.: 7.659 min
Delta R.T.: -0.002 min
Response: 332291
Conc: 0.94 ng/ml




#40 AR-1262-5

R.T.: 8.199 min
Delta R.T.: 0.006 min
Response: 84641
Conc: 0.23 ng/ml



#40 AR-1262-5


R.T.: 8.091 min
Delta R.T.: -0.063 min
Response: 64874
Conc: 0.45 ng/ml



#41 AR-1268-1


R.T.: 7.640 min
Delta R.T.: 0.007 min
Response: 154424
Conc: 0.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#41 AR-1268-1

R.T.: 7.619 min
Delta R.T.: 0.022 min
Response: 364595
Conc: 0.64 ng/ml




#42 AR-1268-2

R.T.: 7.690 min
Delta R.T.: -0.007 min
Response: 156071
Conc: 0.13 ng/ml



#42 AR-1268-2


R.T.: 7.659 min
Delta R.T.: -0.003 min
Response: 332291
Conc: 0.63 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 7.905 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#43 AR-1268-3

R.T.: 7.860 min
Delta R.T.: -0.010 min
Response: 72406
Conc: 0.17 ng/ml




#44 AR-1268-4

R.T.: 8.199 min
Delta R.T.: 0.006 min
Response: 84641
Conc: 0.20 ng/ml



#44 AR-1268-4


R.T.: 8.091 min
Delta R.T.: -0.063 min
Response: 64874
Conc: 0.40 ng/ml



#45 AR-1268-5

R.T.: 8.500 min
Delta R.T.: 0.007 min
Response: 116273
Conc: 0.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE



#45 AR-1268-5

R.T.: 8.450 min
Delta R.T.: 0.002 min
Response: 55286
Conc: 0.05 ng/ml