

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022822\
 Data File : P0084899.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Feb 2022 11:31
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 13:38:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 13:36:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.594	14394617	2280427	50.000	50.000
2) SA Decachlor...	10.428	8.620	9024230	1840572	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.697	4.685	4693414	806441	500.000	500.000
4) L1 AR-1016-2	5.720	4.704	6704457	1143787	500.000	500.000
5) L1 AR-1016-3	5.783	4.880	4031331	628108	500.000	500.000
6) L1 AR-1016-4	5.883	4.923	3364380	517086	500.000	500.000
7) L1 AR-1016-5	6.182	5.136	3194279	651435	500.000	500.000
31) L7 AR-1260-1	7.325	6.174	5672432	1221001	500.000	500.000
32) L7 AR-1260-2	7.585	6.363	6418887	1464665	500.000	500.000
33) L7 AR-1260-3	7.950	6.516	3960602	1367612	500.000	500.000
34) L7 AR-1260-4	8.180	6.991	4796191	1007010	500.000	500.000
35) L7 AR-1260-5	8.513	7.234	9341061	2354684	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022822\
Data File : P0084899.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Feb 2022 11:31
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 13:38:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 28 13:36:52 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 14394617
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: 0.000 min
Response: 2280427
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.428 min
Delta R.T.: 0.000 min
Response: 9024230
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.620 min
Delta R.T.: 0.000 min
Response: 1840572
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 4693414
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.685 min
Delta R.T.: 0.000 min
Response: 806441
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 6704457
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 1143787
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 4031331
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.880 min
Delta R.T.: 0.000 min
Response: 628108
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.883 min
Delta R.T.: 0.000 min
Response: 3364380
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 517086
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 3194279
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 651435
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.325 min
Delta R.T.: 0.000 min
Response: 5672432
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1221001
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.585 min
Delta R.T.: 0.000 min
Response: 6418887
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.363 min
Delta R.T.: 0.000 min
Response: 1464665
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 3960602
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 1367612
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.180 min
Delta R.T.: 0.000 min
Response: 4796191
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 6.991 min
Delta R.T.: 0.000 min
Response: 1007010
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.513 min
Delta R.T.: 0.000 min
Response: 9341061
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.234 min
Delta R.T.: 0.000 min
Response: 2354684
Conc: 500.00 ng/ml