

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022924\
 Data File : P0102222.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Feb 2024 10:45
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:12:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.276	3.390	211.4E6	141.8E6	45.575	43.640
2)	SA Decachlor...	9.823	8.262	118.5E6	83034752	46.545	40.137
Target Compounds							
21)	L5 AR-1248-1	5.421	4.443	37548971	28909403	502.398	524.330
22)	L5 AR-1248-2	5.690	4.674	57820504	41094836	498.422	489.618
23)	L5 AR-1248-3	5.891	4.714	61007068	44230132	500.530	503.739
24)	L5 AR-1248-4	6.289	4.882	61124302	48681820	506.527	483.396
25)	L5 AR-1248-5	6.327	5.267	61814718	43958976	517.942	502.762



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022924\
Data File : PO102222.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Feb 2024 10:45
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 00:12:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.276 min
Delta R.T.: -0.006 min
Response: 211357801
Conc: 45.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.390 min
Delta R.T.: -0.002 min
Response: 141804518
Conc: 43.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.823 min
Delta R.T.: -0.018 min
Response: 118522205
Conc: 46.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.262 min
Delta R.T.: -0.010 min
Response: 83034752
Conc: 40.14 ng/ml



#21 AR-1248-1


R.T.: 5.421 min
Delta R.T.: -0.006 min
Response: 37548971
Conc: 502.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.443 min
Delta R.T.: -0.004 min
Response: 28909403
Conc: 524.33 ng/ml




#22 AR-1248-2

R.T.: 5.690 min
Delta R.T.: -0.007 min
Response: 57820504
Conc: 498.42 ng/ml



#22 AR-1248-2


R.T.: 4.674 min
Delta R.T.: -0.004 min
Response: 41094836
Conc: 489.62 ng/ml



#23 AR-1248-3


R.T.: 5.891 min
Delta R.T.: -0.008 min
Response: 61007068
Conc: 500.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.714 min
Delta R.T.: -0.004 min
Response: 44230132
Conc: 503.74 ng/ml




#24 AR-1248-4

R.T.: 6.289 min
Delta R.T.: -0.008 min
Response: 61124302
Conc: 506.53 ng/ml



#24 AR-1248-4


R.T.: 4.882 min
Delta R.T.: -0.005 min
Response: 48681820
Conc: 483.40 ng/ml



#25 AR-1248-5

R.T.: 6.327 min
Delta R.T.: -0.008 min
Response: 61814718
Conc: 517.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.267 min
Delta R.T.: -0.005 min
Response: 43958976
Conc: 502.76 ng/ml