

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030122\
 Data File : P0084940.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Mar 2022 14:21
 Operator : YP\AJ
 Sample : N1701-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 00:16:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 18:40:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.508	3.596	4344794	849851	15.540	18.898
2)	SA Decachlor...	10.431	8.621	2046848	446170	11.677	12.142
Target Compounds							
3)	L1 AR-1016-1	5.679	0.000	13964	0	1.491	N.D. #
4)	L1 AR-1016-2	5.717	0.000	12033	0	0.902	N.D. #
5)	L1 AR-1016-3	5.776	0.000	6430	0	0.805	N.D. #
6)	L1 AR-1016-4	5.882	0.000	35440	0	5.276	N.D. #
7)	L1 AR-1016-5	6.194	0.000	16348	0	2.586	N.D. #
8)	L2 AR-1221-1	4.721	3.757f	1166833	-1120	364.689	N.D. #
9)	L2 AR-1221-2	4.820	3.880	57134	98496	23.695	224.126 #
10)	L2 AR-1221-3	4.901	4.017f	28001	3449	3.845	2.590 #
11)	L3 AR-1232-1	4.869	4.017f	5582	3449	0.912	3.100 #
12)	L3 AR-1232-2	5.432	0.000	12532	0	4.212	N.D. #
13)	L3 AR-1232-3	5.717	0.000	12033	0	2.122	N.D. #
14)	L3 AR-1232-4	5.882	0.000	35440	0	12.409	N.D. #
15)	L3 AR-1232-5	5.955	0.000	5122	0	2.267	N.D. #
16)	L4 AR-1242-1	5.679	0.000	13964	0	1.881	N.D. #
17)	L4 AR-1242-2	5.717	0.000	12033	0	1.136	N.D. #
18)	L4 AR-1242-3	5.776	0.000	6430	0	1.000	N.D. #
19)	L4 AR-1242-4	5.882	0.000	35440	0	6.536	N.D. #
20)	L4 AR-1242-5	6.635	0.000	2171	0	0.413	N.D. #
21)	L5 AR-1248-1	5.679	0.000	13964	0	2.574	N.D. #
22)	L5 AR-1248-2	5.955	0.000	5122	0	0.666	N.D. #
23)	L5 AR-1248-3	6.194	0.000	16348	0	1.903	N.D. #
24)	L5 AR-1248-4	6.596	0.000	27884	0	2.993	N.D. #
25)	L5 AR-1248-5	6.635	0.000	2171	0	0.243	N.D. #
26)	L6 AR-1254-1	6.558	0.000	3917	0	0.399	N.D. #
27)	L6 AR-1254-2	6.776	5.690f	11125	9184	0.750	3.894 #
28)	L6 AR-1254-3	7.159	0.000	25101	0	1.658	N.D. #
29)	L6 AR-1254-4	7.448	0.000	16923	0	1.578	N.D. #
30)	L6 AR-1254-5	7.879	6.691	23602	1913	2.004	0.565 #
31)	L7 AR-1260-1	7.317	6.173	1216	4187	0.107	1.687 #
32)	L7 AR-1260-2	7.581	6.372	10041	4887	0.781	1.651 #
33)	L7 AR-1260-3	7.955	0.000	73774	0	9.238	N.D. #
34)	L7 AR-1260-4	8.158	0.000	13421	0	1.411	N.D. #
35)	L7 AR-1260-5	8.543	7.283f	2975758	1702	161.864	0.366 #
36)	L8 AR-1262-1	7.955	6.691	73774	1913	8.161	1.585 #
37)	L8 AR-1262-2	8.543	0.000	2975758	0	186.536	N.D. #
38)	L8 AR-1262-3	8.839	7.526	35986	3331	3.432	2.143 #
39)	L8 AR-1262-4	8.950	7.642f	5893	1627	1.166	0.574 #
40)	L8 AR-1262-5	9.631	8.044f	64027	2115	11.708	1.598 #
41)	L9 AR-1268-1	8.839	7.526	35986	3331	1.288	0.488 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030122\
 Data File : P0084940.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Mar 2022 14:21
 Operator : YP\AJ
 Sample : N1701-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 00:16:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 18:40:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.950	7.642f	5893	1627	0.235	0.268
43) L9	AR-1268-3	9.184	7.741f	1858	305	0.087	0.061 #
44) L9	AR-1268-4	9.631	8.044f	64027	2115	6.977	0.927 #
45) L9	AR-1268-5	10.058	8.367	4238170	4013	60.462	0.273 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030122\
Data File : P0084940.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Mar 2022 14:21
Operator : YP\AJ
Sample : N1701-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 00:16:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 28 18:40:06 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: 0.002 min
Response: 4344794
Conc: 15.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.596 min
Delta R.T.: 0.002 min
Response: 849851
Conc: 18.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: 0.005 min
Response: 2046848
Conc: 11.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.621 min
Delta R.T.: 0.002 min
Response: 446170
Conc: 12.14 ng/ml



#3 AR-1016-1


R.T.: 5.679 min
Delta R.T.: -0.017 min
Response: 13964
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.685 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.717 min
Delta R.T.: -0.002 min
Response: 12033
Conc: 0.90 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.703 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.776 min
Delta R.T.: -0.007 min
Response: 6430
Conc: 0.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.879 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.882 min
Delta R.T.: 0.000 min
Response: 35440
Conc: 5.28 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.923 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.194 min
Delta R.T.: 0.013 min
Response: 16348
Conc: 2.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.136 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.721 min
Delta R.T.: 0.000 min
Response: 1166833
Conc: 364.69 ng/ml



#8 AR-1221-1


R.T.: 3.757 min
Delta R.T.: -0.054 min
Response: -1120
Conc: N.D.



#9 AR-1221-2


R.T.: 4.820 min
Delta R.T.: 0.013 min
Response: 57134
Conc: 23.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.880 min
Delta R.T.: -0.016 min
Response: 98496
Conc: 224.13 ng/ml




#10 AR-1221-3

R.T.: 4.901 min
Delta R.T.: 0.016 min
Response: 28001
Conc: 3.85 ng/ml



#10 AR-1221-3


R.T.: 4.017 min
Delta R.T.: 0.045 min
Response: 3449
Conc: 2.59 ng/ml



#11 AR-1232-1


R.T.: 4.869 min
Delta R.T.: -0.015 min
Response: 5582
Conc: 0.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.017 min
Delta R.T.: 0.046 min
Response: 3449
Conc: 3.10 ng/ml




#12 AR-1232-2

R.T.: 5.432 min
Delta R.T.: 0.007 min
Response: 12532
Conc: 4.21 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.704 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.717 min
Delta R.T.: -0.002 min
Response: 12033
Conc: 2.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.879 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.882 min
Delta R.T.: -0.001 min
Response: 35440
Conc: 12.41 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.964 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.955 min
Delta R.T.: -0.020 min
Response: 5122
Conc: 2.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.135 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.679 min
Delta R.T.: -0.018 min
Response: 13964
Conc: 1.88 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.686 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.717 min
Delta R.T.: -0.003 min
Response: 12033
Conc: 1.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.705 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.776 min
Delta R.T.: -0.007 min
Response: 6430
Conc: 1.00 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.881 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.882 min
Delta R.T.: -0.001 min
Response: 35440
Conc: 6.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.635 min
Delta R.T.: 0.003 min
Response: 2171
Conc: 0.41 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T. : 5.490 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.679 min
Delta R.T.: -0.018 min
Response: 13964
Conc: 2.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.685 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.955 min
Delta R.T.: -0.021 min
Response: 5122
Conc: 0.67 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.194 min
Delta R.T.: 0.011 min
Response: 16348
Conc: 1.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.964 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.596 min
Delta R.T.: 0.003 min
Response: 27884
Conc: 2.99 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.136 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.635 min
Delta R.T.: 0.003 min
Response: 2171
Conc: 0.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.531 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.558 min
Delta R.T.: -0.009 min
Response: 3917
Conc: 0.40 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.490 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.776 min
Delta R.T.: -0.014 min
Response: 11125
Conc: 0.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.690 min
Delta R.T.: 0.051 min
Response: 9184
Conc: 3.89 ng/ml




#28 AR-1254-3

R.T.: 7.159 min
Delta R.T.: -0.001 min
Response: 25101
Conc: 1.66 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.043 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.448 min
Delta R.T.: 0.000 min
Response: 16923
Conc: 1.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.274 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.879 min
Delta R.T.: 0.005 min
Response: 23602
Conc: 2.00 ng/ml



#30 AR-1254-5


R.T.: 6.691 min
Delta R.T.: -0.001 min
Response: 1913
Conc: 0.57 ng/ml



#31 AR-1260-1


R.T.: 7.317 min
Delta R.T.: -0.008 min
Response: 1216
Conc: 0.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 4187
Conc: 1.69 ng/ml




#32 AR-1260-2

R.T.: 7.581 min
Delta R.T.: -0.003 min
Response: 10041
Conc: 0.78 ng/ml



#32 AR-1260-2


R.T.: 6.372 min
Delta R.T.: 0.010 min
Response: 4887
Conc: 1.65 ng/ml



#33 AR-1260-3


R.T.: 7.955 min
Delta R.T.: 0.006 min
Response: 73774
Conc: 9.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.516 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.158 min
Delta R.T.: -0.021 min
Response: 13421
Conc: 1.41 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.990 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.543 min
Delta R.T.: 0.033 min
Response: 2975758
Conc: 161.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: 0.049 min
Response: 1702
Conc: 0.37 ng/ml




#36 AR-1262-1

R.T.: 7.955 min
Delta R.T.: 0.004 min
Response: 73774
Conc: 8.16 ng/ml



#36 AR-1262-1


R.T.: 6.691 min
Delta R.T.: 0.000 min
Response: 1913
Conc: 1.59 ng/ml



#37 AR-1262-2


R.T.: 8.543 min
Delta R.T.: 0.029 min
Response: 2975758
Conc: 186.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.992 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.839 min
Delta R.T.: -0.004 min
Response: 35986
Conc: 3.43 ng/ml



#38 AR-1262-3


R.T.: 7.526 min
Delta R.T.: 0.004 min
Response: 3331
Conc: 2.14 ng/ml



#39 AR-1262-4


R.T.: 8.950 min
Delta R.T.: 0.016 min
Response: 5893
Conc: 1.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.642 min
Delta R.T.: 0.057 min
Response: 1627
Conc: 0.57 ng/ml




#40 AR-1262-5

R.T.: 9.631 min
Delta R.T.: 0.003 min
Response: 64027
Conc: 11.71 ng/ml



#40 AR-1262-5


R.T.: 8.044 min
Delta R.T.: -0.037 min
Response: 2115
Conc: 1.60 ng/ml



#41 AR-1268-1


R.T.: 8.839 min
Delta R.T.: -0.001 min
Response: 35986
Conc: 1.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.526 min
Delta R.T.: 0.005 min
Response: 3331
Conc: 0.49 ng/ml




#42 AR-1268-2

R.T.: 8.950 min
Delta R.T.: 0.012 min
Response: 5893
Conc: 0.24 ng/ml



#42 AR-1268-2


R.T.: 7.642 min
Delta R.T.: 0.056 min
Response: 1627
Conc: 0.27 ng/ml



#43 AR-1268-3


R.T.: 9.184 min
Delta R.T.: 0.005 min
Response: 1858
Conc: 0.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.741 min
Delta R.T.: -0.052 min
Response: 305
Conc: 0.06 ng/ml




#44 AR-1268-4

R.T.: 9.631 min
Delta R.T.: 0.003 min
Response: 64027
Conc: 6.98 ng/ml



#44 AR-1268-4


R.T.: 8.044 min
Delta R.T.: -0.036 min
Response: 2115
Conc: 0.93 ng/ml



#45 AR-1268-5

R.T.: 10.058 min
Delta R.T.: -0.010 min
Response: 4238170
Conc: 60.46 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.367 min
Delta R.T.: -0.002 min
Response: 4013
Conc: 0.27 ng/ml