

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030122\
 Data File : P0084957.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Mar 2022 19:32
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 00:29:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 18:40:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.508	3.595	11909141	2207768	42.594	49.093
2)	SA Decachlor...	10.432	8.622	7173377	1607916	40.924	43.758
Target Compounds							
3)	L1 AR-1016-1	5.697	4.687	3223290	768325	344.206	476.007 #
4)	L1 AR-1016-2	5.720	4.705	4642539	1086769	348.050	478.330 #
5)	L1 AR-1016-3	5.784	4.881	2861552	590759	358.144	462.363 #
6)	L1 AR-1016-4	5.884	4.924	2237635	486554	333.136	462.772 #
7)	L1 AR-1016-5	6.183	5.137	2114970	594110	334.536	448.300 #
31)	L7 AR-1260-1	7.326	6.175	3576658	1081362	313.434	435.723 #
32)	L7 AR-1260-2	7.586	6.365	4073527	1299351	316.799	438.978 #
33)	L7 AR-1260-3	7.951	6.518	2654605	1206884	332.415	441.287 #
34)	L7 AR-1260-4	8.182	6.992	3197415	874209	336.160	436.078 #
35)	L7 AR-1260-5	8.515	7.236	6689998	2019836	363.896	434.350



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030122\
Data File : P0084957.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Mar 2022 19:32
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 00:29:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 28 18:40:06 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: 0.002 min
Response: 11909141
Conc: 42.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.595 min
Delta R.T.: 0.002 min
Response: 2207768
Conc: 49.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: 0.005 min
Response: 7173377
Conc: 40.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.622 min
Delta R.T.: 0.004 min
Response: 1607916
Conc: 43.76 ng/ml



#3 AR-1016-1


R.T.: 5.697 min
Delta R.T.: 0.002 min
Response: 3223290
Conc: 344.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.687 min
Delta R.T.: 0.001 min
Response: 768325
Conc: 476.01 ng/ml




#4 AR-1016-2

R.T.: 5.720 min
Delta R.T.: 0.001 min
Response: 4642539
Conc: 348.05 ng/ml



#4 AR-1016-2


R.T.: 4.705 min
Delta R.T.: 0.002 min
Response: 1086769
Conc: 478.33 ng/ml



#5 AR-1016-3


R.T.: 5.784 min
Delta R.T.: 0.001 min
Response: 2861552
Conc: 358.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.881 min
Delta R.T.: 0.002 min
Response: 590759
Conc: 462.36 ng/ml




#6 AR-1016-4

R.T.: 5.884 min
Delta R.T.: 0.002 min
Response: 2237635
Conc: 333.14 ng/ml



#6 AR-1016-4


R.T.: 4.924 min
Delta R.T.: 0.001 min
Response: 486554
Conc: 462.77 ng/ml



#7 AR-1016-5


R.T.: 6.183 min
Delta R.T.: 0.002 min
Response: 2114970
Conc: 334.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.137 min
Delta R.T.: 0.001 min
Response: 594110
Conc: 448.30 ng/ml




#31 AR-1260-1

R.T.: 7.326 min
Delta R.T.: 0.002 min
Response: 3576658
Conc: 313.43 ng/ml



#31 AR-1260-1


R.T.: 6.175 min
Delta R.T.: 0.003 min
Response: 1081362
Conc: 435.72 ng/ml



#32 AR-1260-2


R.T.: 7.586 min
Delta R.T.: 0.002 min
Response: 4073527
Conc: 316.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.365 min
Delta R.T.: 0.003 min
Response: 1299351
Conc: 438.98 ng/ml




#33 AR-1260-3

R.T.: 7.951 min
Delta R.T.: 0.002 min
Response: 2654605
Conc: 332.41 ng/ml



#33 AR-1260-3


R.T.: 6.518 min
Delta R.T.: 0.002 min
Response: 1206884
Conc: 441.29 ng/ml



#34 AR-1260-4


R.T.: 8.182 min
Delta R.T.: 0.003 min
Response: 3197415
Conc: 336.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.992 min
Delta R.T.: 0.002 min
Response: 874209
Conc: 436.08 ng/ml



#35 AR-1260-5

R.T.: 8.515 min
Delta R.T.: 0.004 min
Response: 6689998
Conc: 363.90 ng/ml



#35 AR-1260-5

R.T.: 7.236 min
Delta R.T.: 0.003 min
Response: 2019836
Conc: 434.35 ng/ml