

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0030222\
 Data File : P0084985.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Mar 2022 10:54
 Operator : YP\AJ
 Sample : N1715-04 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 03:54:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 18:40:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.508	3.595	303191	75028	1.084	1.668 #
2)	SA Decachlor...	10.434	8.623	245193	54378	1.399	1.480
Target Compounds							
3)	L1 AR-1016-1	5.679	0.000	12221	0	1.305	N.D. #
4)	L1 AR-1016-2	5.757f	0.000	10816	0	0.811	N.D. #
5)	L1 AR-1016-3	5.757	0.000	10816	0	1.354	N.D. #
6)	L1 AR-1016-4	5.883	0.000	4038	0	0.601	N.D. #
7)	L1 AR-1016-5	6.185	5.152	15970	16495	2.526	12.447 #
8)	L2 AR-1221-1	4.727	3.757f	3547	21218	1.109	36.390 #
9)	L2 AR-1221-2	4.789	0.000	17118	0	7.099	N.D. #
10)	L2 AR-1221-3	4.887	0.000	4298	0	0.590	N.D. #
11)	L3 AR-1232-1	4.881	0.000	5595	0	0.914	N.D. #
12)	L3 AR-1232-2	5.423	0.000	8580	0	2.884	N.D. #
13)	L3 AR-1232-3	5.757f	0.000	10816	0	1.908	N.D. #
14)	L3 AR-1232-4	5.883	0.000	4038	0	1.414	N.D. #
15)	L3 AR-1232-5	5.985	5.152	2748	16495	1.216	29.745 #
16)	L4 AR-1242-1	5.679	0.000	12221	0	1.646	N.D. #
17)	L4 AR-1242-2	5.757f	0.000	10816	0	1.021	N.D. #
18)	L4 AR-1242-3	5.757	0.000	10816	0	1.683	N.D. #
19)	L4 AR-1242-4	5.883	0.000	4038	0	0.745	N.D. #
20)	L4 AR-1242-5	6.640	5.516	39624	338584	7.534	273.194 #
21)	L5 AR-1248-1	5.679	0.000	12221	0	2.252	N.D. #
22)	L5 AR-1248-2	5.985	0.000	2748	0	0.357	N.D. #
23)	L5 AR-1248-3	6.185	0.000	15970	0	1.859	N.D. #
24)	L5 AR-1248-4	6.611	5.152	26963	16495	2.894	9.496 #
25)	L5 AR-1248-5	6.640	5.516	39624	338584	4.432	201.890 #
26)	L6 AR-1254-1	6.568	5.516	1565709	338584	159.629	126.564
27)	L6 AR-1254-2	6.793	0.000	12560	0	0.846	N.D. #
28)	L6 AR-1254-3	7.162	0.000	235774	0	15.572	N.D. #
29)	L6 AR-1254-4	7.456	6.343f	62802	43639	5.855	18.618 #
30)	L6 AR-1254-5	7.877	0.000	46069	0	3.912	N.D. #
31)	L7 AR-1260-1	7.339	6.170	125170	3422	10.969	1.379 #
32)	L7 AR-1260-2	7.582	6.343	80168	43639	6.235	14.743 #
33)	L7 AR-1260-3	7.951	6.575f	42155	21986	5.279	8.039 #
34)	L7 AR-1260-4	8.198	0.000	24732	0	2.600	N.D. #
35)	L7 AR-1260-5	8.527	0.000	14373	0	0.782	N.D. #
36)	L8 AR-1262-1	7.951	0.000	42155	0	4.663	N.D. #
37)	L8 AR-1262-2	8.527	0.000	14373	0	0.901	N.D. #
38)	L8 AR-1262-3	8.836	7.535	5466	546203	0.521	351.375 #
39)	L8 AR-1262-4	8.947	7.535f	4223	546203	0.836	192.790 #
40)	L8 AR-1262-5	9.653	8.125f	1124	24429	0.206	18.456 #
41)	L9 AR-1268-1	8.836	7.535	5466	546203	0.196	80.068 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030222\
 Data File : P0084985.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Mar 2022 10:54
 Operator : YP\AJ
 Sample : N1715-04 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 03:54:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 28 18:40:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.947	7.535f	4223	546203	0.169	90.121 #
43) L9	AR-1268-3	9.173	0.000	7511	0	0.352	N.D. #
44) L9	AR-1268-4	9.653	8.125f	1124	24429	0.122	10.703 #
45) L9	AR-1268-5	10.066	8.383	5662	1458	0.081	0.099



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030222\
Data File : P0084985.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Mar 2022 10:54
Operator : YP\AJ
Sample : N1715-04 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 03:54:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022822.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 28 18:40:06 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: 0.002 min
Response: 303191
Conc: 1.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.595 min
Delta R.T.: 0.001 min
Response: 75028
Conc: 1.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.007 min
Response: 245193
Conc: 1.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.623 min
Delta R.T.: 0.004 min
Response: 54378
Conc: 1.48 ng/ml



#3 AR-1016-1


R.T.: 5.679 min
Delta R.T.: -0.016 min
Response: 12221
Conc: 1.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.685 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.757 min
Delta R.T.: 0.037 min
Response: 10816
Conc: 0.81 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.703 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.757 min
Delta R.T.: -0.026 min
Response: 10816
Conc: 1.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.879 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.883 min
Delta R.T.: 0.000 min
Response: 4038
Conc: 0.60 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.923 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.185 min
Delta R.T.: 0.003 min
Response: 15970
Conc: 2.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.152 min
Delta R.T.: 0.016 min
Response: 16495
Conc: 12.45 ng/ml




#8 AR-1221-1

R.T.: 4.727 min
Delta R.T.: 0.007 min
Response: 3547
Conc: 1.11 ng/ml



#8 AR-1221-1


R.T.: 3.757 min
Delta R.T.: -0.054 min
Response: 21218
Conc: 36.39 ng/ml



#9 AR-1221-2


R.T.: 4.789 min
Delta R.T.: -0.018 min
Response: 17118
Conc: 7.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.896 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.887 min
Delta R.T.: 0.002 min
Response: 4298
Conc: 0.59 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.973 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.881 min
Delta R.T.: -0.003 min
Response: 5595
Conc: 0.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.972 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.423 min
Delta R.T.: -0.002 min
Response: 8580
Conc: 2.88 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.704 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.757 min
Delta R.T.: 0.037 min
Response: 10816
Conc: 1.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.879 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.883 min
Delta R.T.: 0.000 min
Response: 4038
Conc: 1.41 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.964 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.985 min
Delta R.T.: 0.010 min
Response: 2748
Conc: 1.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.152 min
Delta R.T.: 0.017 min
Response: 16495
Conc: 29.74 ng/ml




#16 AR-1242-1

R.T.: 5.679 min
Delta R.T.: -0.018 min
Response: 12221
Conc: 1.65 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.686 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.757 min
Delta R.T.: 0.037 min
Response: 10816
Conc: 1.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.705 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.757 min
Delta R.T.: -0.026 min
Response: 10816
Conc: 1.68 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.881 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.883 min
Delta R.T.: 0.000 min
Response: 4038
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.640 min
Delta R.T.: 0.007 min
Response: 39624
Conc: 7.53 ng/ml



#20 AR-1242-5


R.T.: 5.516 min
Delta R.T.: 0.026 min
Response: 338584
Conc: 273.19 ng/ml



#21 AR-1248-1


R.T.: 5.679 min
Delta R.T.: -0.018 min
Response: 12221
Conc: 2.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.685 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.985 min
Delta R.T.: 0.009 min
Response: 2748
Conc: 0.36 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.185 min
Delta R.T.: 0.001 min
Response: 15970
Conc: 1.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.964 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.611 min
Delta R.T.: 0.018 min
Response: 26963
Conc: 2.89 ng/ml



#24 AR-1248-4


R.T.: 5.152 min
Delta R.T.: 0.016 min
Response: 16495
Conc: 9.50 ng/ml



#25 AR-1248-5


R.T.: 6.640 min
Delta R.T.: 0.008 min
Response: 39624
Conc: 4.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.516 min
Delta R.T.: -0.015 min
Response: 338584
Conc: 201.89 ng/ml




#26 AR-1254-1

R.T.: 6.568 min
Delta R.T.: 0.002 min
Response: 1565709
Conc: 159.63 ng/ml



#26 AR-1254-1


R.T.: 5.516 min
Delta R.T.: 0.025 min
Response: 338584
Conc: 126.56 ng/ml



#27 AR-1254-2


R.T.: 6.793 min
Delta R.T.: 0.003 min
Response: 12560
Conc: 0.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.639 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.162 min
Delta R.T.: 0.002 min
Response: 235774
Conc: 15.57 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.043 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.456 min
Delta R.T.: 0.008 min
Response: 62802
Conc: 5.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.343 min
Delta R.T.: 0.069 min
Response: 43639
Conc: 18.62 ng/ml




#30 AR-1254-5

R.T.: 7.877 min
Delta R.T.: 0.004 min
Response: 46069
Conc: 3.91 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.693 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.339 min
Delta R.T.: 0.014 min
Response: 125170
Conc: 10.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.170 min
Delta R.T.: -0.003 min
Response: 3422
Conc: 1.38 ng/ml




#32 AR-1260-2

R.T.: 7.582 min
Delta R.T.: -0.002 min
Response: 80168
Conc: 6.23 ng/ml



#32 AR-1260-2


R.T.: 6.343 min
Delta R.T.: -0.020 min
Response: 43639
Conc: 14.74 ng/ml



#33 AR-1260-3


R.T.: 7.951 min
Delta R.T.: 0.002 min
Response: 42155
Conc: 5.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.575 min
Delta R.T.: 0.059 min
Response: 21986
Conc: 8.04 ng/ml




#34 AR-1260-4

R.T.: 8.198 min
Delta R.T.: 0.019 min
Response: 24732
Conc: 2.60 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.990 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.527 min
Delta R.T.: 0.016 min
Response: 14373
Conc: 0.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.233 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.951 min
Delta R.T.: 0.000 min
Response: 42155
Conc: 4.66 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.692 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.527 min
Delta R.T.: 0.012 min
Response: 14373
Conc: 0.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.992 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.836 min
Delta R.T.: -0.007 min
Response: 5466
Conc: 0.52 ng/ml



#38 AR-1262-3


R.T.: 7.535 min
Delta R.T.: 0.014 min
Response: 546203
Conc: 351.37 ng/ml



#39 AR-1262-4


R.T.: 8.947 min
Delta R.T.: 0.013 min
Response: 4223
Conc: 0.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.535 min
Delta R.T.: -0.050 min
Response: 546203
Conc: 192.79 ng/ml




#40 AR-1262-5

R.T.: 9.653 min
Delta R.T.: 0.025 min
Response: 1124
Conc: 0.21 ng/ml



#40 AR-1262-5


R.T.: 8.125 min
Delta R.T.: 0.044 min
Response: 24429
Conc: 18.46 ng/ml



#41 AR-1268-1


R.T.: 8.836 min
Delta R.T.: -0.004 min
Response: 5466
Conc: 0.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.535 min
Delta R.T.: 0.014 min
Response: 546203
Conc: 80.07 ng/ml




#42 AR-1268-2

R.T.: 8.947 min
Delta R.T.: 0.009 min
Response: 4223
Conc: 0.17 ng/ml



#42 AR-1268-2


R.T.: 7.535 min
Delta R.T.: -0.050 min
Response: 546203
Conc: 90.12 ng/ml



#43 AR-1268-3


R.T.: 9.173 min
Delta R.T.: -0.007 min
Response: 7511
Conc: 0.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.793 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: 0.025 min
Response: 1124
Conc: 0.12 ng/ml



#44 AR-1268-4


R.T.: 8.125 min
Delta R.T.: 0.044 min
Response: 24429
Conc: 10.70 ng/ml



#45 AR-1268-5

R.T.: 10.066 min
Delta R.T.: -0.001 min
Response: 5662
Conc: 0.08 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.383 min
Delta R.T.: 0.014 min
Response: 1458
Conc: 0.10 ng/ml