

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030223\
 Data File : P0092891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2023 14:52
 Operator : YP/AJ
 Sample : 01744-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 72-11939

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/03/2023
 Supervised By :Ankita Jodhani 03/03/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 15:52:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.638	53822336	21455614	17.924	16.832
2) SA Decachlor...	10.276	8.676	37105651	17139224	19.291m	14.707m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030223\
Data File : PO092891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2023 14:52
Operator : YP/AJ
Sample : 01744-01
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
72-11939


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/03/2023
Supervised By :Ankita Jodhani 03/03/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 15:52:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.433 min
Delta R.T.: 0.000 min
Response: 53822336
Conc: 17.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
72-11939


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/03/2023
Supervised By :Ankita Jodhani 03/03/2023




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 21455614
Conc: 16.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.276 min
Delta R.T.: 0.020 min
Response: 37105651
Conc: 19.29 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.676 min
Delta R.T.: 0.004 min
Response: 17139224
Conc: 14.71 ng/ml m