

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030321\
 Data File : P0075843.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Mar 2021 17:22
 Operator : AJ/MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 18:06:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 27 03:46:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.938	3.953	1550815	1169940	25.370	26.482
2)	SA Decachlor...	10.956	9.222	1013179	952615	22.266	21.952
Target Compounds							
9)	L2 AR-1221-2	5.291	4.302	19116	14379	35.971	34.347
20)	L4 AR-1242-5	7.233	6.069	23025	11631	18.265	8.278 #
24)	L5 AR-1248-4	7.204	0.000	99622	0	41.386	N.D. #
25)	L5 AR-1248-5	7.233	0.000	23025	0	9.986	N.D. #
26)	L6 AR-1254-1	7.175	6.069	131568	11631	55.945	4.105 #
27)	L6 AR-1254-2	7.405	6.228	18688	14468	5.203	5.842
28)	L6 AR-1254-3	7.789	6.646	39937	155559	10.517	38.785 #
29)	L6 AR-1254-4	8.083	6.884	13240	51748	4.772	19.692 #
30)	L6 AR-1254-5	8.514	7.311	41574	69661	14.840	19.840 #
31)	L7 AR-1260-1	7.954	6.784	13489	88215	5.502	35.097 #
32)	L7 AR-1260-2	8.221	6.979	23037	28776	7.726	9.464
33)	L7 AR-1260-3	0.000	7.130	0	49664	N.D.	17.877 #
34)	L7 AR-1260-4	8.811	0.000	35117	0	13.506	N.D. #
35)	L7 AR-1260-5	9.147	7.836f	11543	4231	2.212	0.744 #
36)	L8 AR-1262-1	0.000	7.311	0	69661	N.D.	33.258 #
37)	L8 AR-1262-2	9.147	7.836f	11543	4231	1.763	0.610 #
39)	L8 AR-1262-4	0.000	8.205	0	15105	N.D.	2.998 #
42)	L9 AR-1268-2	0.000	8.205	0	15105	N.D.	2.244 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030321\
Data File : PO075843.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Mar 2021 17:22
Operator : AJ/MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 18:06:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022621.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 27 03:46:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.938 min
Delta R.T.: 0.002 min
Response: 1550815
Conc: 25.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.953 min
Delta R.T.: -0.002 min
Response: 1169940
Conc: 26.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.956 min
Delta R.T.: 0.011 min
Response: 1013179
Conc: 22.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.222 min
Delta R.T.: -0.005 min
Response: 952615
Conc: 21.95 ng/ml



#9 AR-1221-2


R.T.: 5.291 min
Delta R.T.: 0.004 min
Response: 19116
Conc: 35.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.302 min
Delta R.T.: -0.008 min
Response: 14379
Conc: 34.35 ng/ml




#20 AR-1242-5

R.T.: 7.233 min
Delta R.T.: -0.009 min
Response: 23025
Conc: 18.27 ng/ml



#20 AR-1242-5


R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 11631
Conc: 8.28 ng/ml



#24 AR-1248-4


R.T.: 7.204 min
Delta R.T.: 0.004 min
Response: 99622
Conc: 41.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.693 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 7.233 min
Delta R.T.: -0.007 min
Response: 23025
Conc: 9.99 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 6.113 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 7.175 min
Delta R.T.: 0.004 min
Response: 131568
Conc: 55.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 6.069 min
Delta R.T.: -0.002 min
Response: 11631
Conc: 4.10 ng/ml




#27 AR-1254-2

R.T.: 7.405 min
Delta R.T.: 0.004 min
Response: 18688
Conc: 5.20 ng/ml



#27 AR-1254-2


R.T.: 6.228 min
Delta R.T.: -0.002 min
Response: 14468
Conc: 5.84 ng/ml



#28 AR-1254-3


R.T.: 7.789 min
Delta R.T.: 0.006 min
Response: 39937
Conc: 10.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.646 min
Delta R.T.: -0.002 min
Response: 15559
Conc: 38.78 ng/ml




#29 AR-1254-4

R.T.: 8.083 min
Delta R.T.: 0.003 min
Response: 13240
Conc: 4.77 ng/ml



#29 AR-1254-4


R.T.: 6.884 min
Delta R.T.: -0.002 min
Response: 51748
Conc: 19.69 ng/ml



#30 AR-1254-5


R.T.: 8.514 min
Delta R.T.: 0.006 min
Response: 41574
Conc: 14.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 7.311 min
Delta R.T.: -0.003 min
Response: 69661
Conc: 19.84 ng/ml




#31 AR-1260-1

R.T.: 7.954 min
Delta R.T.: 0.004 min
Response: 13489
Conc: 5.50 ng/ml



#31 AR-1260-1


R.T.: 6.784 min
Delta R.T.: -0.002 min
Response: 88215
Conc: 35.10 ng/ml



#32 AR-1260-2


R.T.: 8.221 min
Delta R.T.: 0.004 min
Response: 23037
Conc: 7.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.979 min
Delta R.T.: -0.004 min
Response: 28776
Conc: 9.46 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.583 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.130 min
Delta R.T.: -0.007 min
Response: 49664
Conc: 17.88 ng/ml



#34 AR-1260-4


R.T.: 8.811 min
Delta R.T.: -0.001 min
Response: 35117
Conc: 13.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.618 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.147 min
Delta R.T.: 0.008 min
Response: 11543
Conc: 2.21 ng/ml



#35 AR-1260-5


R.T.: 7.836 min
Delta R.T.: -0.028 min
Response: 4231
Conc: 0.74 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 8.583 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#36 AR-1262-1

R.T.: 7.311 min
Delta R.T.: -0.003 min
Response: 69661
Conc: 33.26 ng/ml




#37 AR-1262-2

R.T.: 9.147 min
Delta R.T.: 0.008 min
Response: 11543
Conc: 1.76 ng/ml



#37 AR-1262-2


R.T.: 7.836 min
Delta R.T.: -0.028 min
Response: 4231
Conc: 0.61 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.542 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 8.205 min
Delta R.T.: -0.008 min
Response: 15105
Conc: 3.00 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 9.543 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.205 min
Delta R.T.: -0.010 min
Response: 15105
Conc: 2.24 ng/ml