

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030323\
 Data File : P0092923.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2023 09:39
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 16:46:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.637	143.5E6	55056798	47.776	43.193
2) SA Decachlor...	10.266	8.671	117.0E6	50475382	60.839	43.311 #
Target Compounds						
16) L4 AR-1242-1	5.608	4.723	36239815	14760238	459.306	439.539
17) L4 AR-1242-2	5.630	4.741	54082992	21353667	470.157	445.307
18) L4 AR-1242-3	5.693	4.918	35853345	11557015	485.667	443.821
19) L4 AR-1242-4	5.792	5.002	28044701	12734663	506.343	441.632
20) L4 AR-1242-5	6.535	5.526	28527855	14234106	487.049	454.922



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030323\
Data File : PO092923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2023 09:39
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 16:46:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: -0.003 min
Response: 143460210
Conc: 47.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.002 min
Response: 55056798
Conc: 43.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.266 min
Delta R.T.: 0.010 min
Response: 117022510
Conc: 60.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.671 min
Delta R.T.: 0.000 min
Response: 50475382
Conc: 43.31 ng/ml



#16 AR-1242-1


R.T.: 5.608 min
Delta R.T.: -0.001 min
Response: 36239815
Conc: 459.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.723 min
Delta R.T.: -0.002 min
Response: 14760238
Conc: 439.54 ng/ml




#17 AR-1242-2

R.T.: 5.630 min
Delta R.T.: -0.002 min
Response: 54082992
Conc: 470.16 ng/ml



#17 AR-1242-2


R.T.: 4.741 min
Delta R.T.: -0.002 min
Response: 21353667
Conc: 445.31 ng/ml



#18 AR-1242-3


R.T.: 5.693 min
Delta R.T.: -0.001 min
Response: 35853345
Conc: 485.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.918 min
Delta R.T.: -0.002 min
Response: 11557015
Conc: 443.82 ng/ml




#19 AR-1242-4

R.T.: 5.792 min
Delta R.T.: -0.001 min
Response: 28044701
Conc: 506.34 ng/ml



#19 AR-1242-4


R.T.: 5.002 min
Delta R.T.: -0.001 min
Response: 12734663
Conc: 441.63 ng/ml



#20 AR-1242-5

R.T.: 6.535 min
Delta R.T.: 0.000 min
Response: 28527855
Conc: 487.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.526 min
Delta R.T.: -0.001 min
Response: 14234106
Conc: 454.92 ng/ml