

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030325\
 Data File : P0109624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2025 10:36
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 01:06:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 04:40:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.696	3.693	445.7E6	261.0E6	47.086	49.866
2) SA Decachlor...	8.750	8.701	361.1E6	144.6E6	41.980	45.420
Target Compounds						
21) L5 AR-1248-1	4.789	4.776	95135374	52861661	476.847	525.845
22) L5 AR-1248-2	5.027	5.012	130.4E6	73832428	468.533	507.510
23) L5 AR-1248-3	5.242	5.054	159.8E6	78216651	457.438	504.597
24) L5 AR-1248-4	5.597	5.225	227.2E6	91163314	476.228	504.230
25) L5 AR-1248-5	5.639	5.618	161.3E6	89480748	471.020	519.620



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030325\
Data File : PO109624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2025 10:36
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 01:06:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.696 min
Delta R.T.: -0.002 min
Response: 445673422
Conc: 47.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 261009054
Conc: 49.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.750 min
Delta R.T.: -0.007 min
Response: 361110243
Conc: 41.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.701 min
Delta R.T.: -0.006 min
Response: 144629889
Conc: 45.42 ng/ml



#21 AR-1248-1


R.T.: 4.789 min
Delta R.T.: -0.003 min
Response: 95135374
Conc: 476.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.776 min
Delta R.T.: -0.001 min
Response: 52861661
Conc: 525.84 ng/ml




#22 AR-1248-2

R.T.: 5.027 min
Delta R.T.: -0.003 min
Response: 130388450
Conc: 468.53 ng/ml



#22 AR-1248-2


R.T.: 5.012 min
Delta R.T.: -0.002 min
Response: 73832428
Conc: 507.51 ng/ml



#23 AR-1248-3


R.T.: 5.242 min
Delta R.T.: -0.003 min
Response: 159816303
Conc: 457.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.054 min
Delta R.T.: -0.002 min
Response: 78216651
Conc: 504.60 ng/ml




#24 AR-1248-4

R.T.: 5.597 min
Delta R.T.: -0.003 min
Response: 227205671
Conc: 476.23 ng/ml



#24 AR-1248-4


R.T.: 5.225 min
Delta R.T.: -0.002 min
Response: 91163314
Conc: 504.23 ng/ml



#25 AR-1248-5

R.T.: 5.639 min
Delta R.T.: -0.003 min
Response: 161328697
Conc: 471.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.618 min
Delta R.T.: -0.002 min
Response: 89480748
Conc: 519.62 ng/ml