

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030421\
 Data File : P0075852.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Mar 2021 18:01
 Operator : AJ/MA
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 06:39:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 06:37:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.930	3.946	293934	230439	4.869	5.160
2)	SA Decachlor...	10.946	9.227	151912	192573	4.488	5.302
Target Compounds							
3)	L1 AR-1016-1	6.255	5.197	95182	79126	52.236	54.178
4)	L1 AR-1016-2	6.279	5.217	145183	136820	51.562	58.258
5)	L1 AR-1016-3	6.345	5.407	85427	60526	51.333	51.051
6)	L1 AR-1016-4	6.453	5.459	64241	47598	48.581	45.958
7)	L1 AR-1016-5	6.768	5.686	68651	66886	51.699	52.000
31)	L7 AR-1260-1	7.949	6.781	110200	126799	51.049	57.189
32)	L7 AR-1260-2	8.216	6.978	124160	163787	50.642	61.624
33)	L7 AR-1260-3	8.582	7.132	90606	168068	49.046	66.869 #
34)	L7 AR-1260-4	8.813	7.615	157049	106081	72.913	51.735 #
35)	L7 AR-1260-5	9.139	7.862	188699	257074	48.996	53.393



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030421\
Data File : P0075852.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Mar 2021 18:01
Operator : AJ/MA
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 06:39:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 06:37:40 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 293934
Conc: 4.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.946 min
Delta R.T.: 0.000 min
Response: 230439
Conc: 5.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.946 min
Delta R.T.: -0.002 min
Response: 151912
Conc: 4.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.227 min
Delta R.T.: 0.000 min
Response: 192573
Conc: 5.30 ng/ml



#3 AR-1016-1


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 95182
Conc: 52.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 79126
Conc: 54.18 ng/ml




#4 AR-1016-2

R.T.: 6.279 min
Delta R.T.: 0.000 min
Response: 145183
Conc: 51.56 ng/ml



#4 AR-1016-2


R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 136820
Conc: 58.26 ng/ml



#5 AR-1016-3


R.T.: 6.345 min
Delta R.T.: 0.000 min
Response: 85427
Conc: 51.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 5.407 min
Delta R.T.: 0.000 min
Response: 60526
Conc: 51.05 ng/ml




#6 AR-1016-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 64241
Conc: 48.58 ng/ml



#6 AR-1016-4


R.T.: 5.459 min
Delta R.T.: 0.000 min
Response: 47598
Conc: 45.96 ng/ml



#7 AR-1016-5


R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 68651
Conc: 51.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 66886
Conc: 52.00 ng/ml




#31 AR-1260-1

R.T.: 7.949 min
Delta R.T.: 0.000 min
Response: 110200
Conc: 51.05 ng/ml



#31 AR-1260-1


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 126799
Conc: 57.19 ng/ml



#32 AR-1260-2


R.T.: 8.216 min
Delta R.T.: 0.001 min
Response: 124160
Conc: 50.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.978 min
Delta R.T.: 0.000 min
Response: 163787
Conc: 61.62 ng/ml




#33 AR-1260-3

R.T.: 8.582 min
Delta R.T.: 0.000 min
Response: 90606
Conc: 49.05 ng/ml



#33 AR-1260-3


R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 168068
Conc: 66.87 ng/ml



#34 AR-1260-4


R.T.: 8.813 min
Delta R.T.: 0.000 min
Response: 157049
Conc: 72.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 106081
Conc: 51.73 ng/ml



#35 AR-1260-5

R.T.: 9.139 min
Delta R.T.: 0.000 min
Response: 188699
Conc: 49.00 ng/ml



#35 AR-1260-5

R.T.: 7.862 min
Delta R.T.: 0.000 min
Response: 257074
Conc: 53.39 ng/ml