

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030421\
 Data File : P0075871.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Mar 2021 23:27
 Operator : AJ/MA
 Sample : AR1262ICC750
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 07:35:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 07:35:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.928	3.944	4822515	3489208	74.701	73.619
2) SA Decachlor...	10.941	9.224	2594007	2776612	73.143	72.129
Target Compounds						
36) L8 AR-1262-1	8.579	7.309	2093285	1192806	732.154	713.127
37) L8 AR-1262-2	9.136	7.860	3404319	4007551	740.553	729.072
38) L8 AR-1262-3	9.445	8.147	2248040	1642993	739.953	717.495
39) L8 AR-1262-4	9.537	8.210	1789067	2966188	769.244	725.314
40) L8 AR-1262-5	10.201	8.707	1124423	1308644	742.086	727.172



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030421\
Data File : P0075871.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Mar 2021 23:27
Operator : AJ/MA
Sample : AR1262ICC750
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 07:35:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 07:35:04 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.928 min
Delta R.T.: 0.000 min
Response: 4822515
Conc: 74.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.944 min
Delta R.T.: 0.000 min
Response: 3489208
Conc: 73.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.941 min
Delta R.T.: 0.000 min
Response: 2594007
Conc: 73.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.224 min
Delta R.T.: 0.000 min
Response: 2776612
Conc: 72.13 ng/ml



#36 AR-1262-1


R.T.: 8.579 min
Delta R.T.: 0.000 min
Response: 2093285
Conc: 732.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#36 AR-1262-1

R.T.: 7.309 min
Delta R.T.: 0.000 min
Response: 1192806
Conc: 713.13 ng/ml




#37 AR-1262-2

R.T.: 9.136 min
Delta R.T.: 0.000 min
Response: 3404319
Conc: 740.55 ng/ml



#37 AR-1262-2


R.T.: 7.860 min
Delta R.T.: 0.000 min
Response: 4007551
Conc: 729.07 ng/ml



#38 AR-1262-3


R.T.: 9.445 min
Delta R.T.: 0.000 min
Response: 2248040
Conc: 739.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#38 AR-1262-3

R.T.: 8.147 min
Delta R.T.: 0.000 min
Response: 1642993
Conc: 717.50 ng/ml




#39 AR-1262-4

R.T.: 9.537 min
Delta R.T.: 0.000 min
Response: 1789067
Conc: 769.24 ng/ml



#39 AR-1262-4


R.T.: 8.210 min
Delta R.T.: 0.000 min
Response: 2966188
Conc: 725.31 ng/ml



#40 AR-1262-5

R.T.: 10.201 min
Delta R.T.: 0.000 min
Response: 1124423
Conc: 742.09 ng/ml

Instrument :
ECD_O
ClientSampleId :



#40 AR-1262-5

R.T.: 8.707 min
Delta R.T.: 0.000 min
Response: 1308644
Conc: 727.17 ng/ml