

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030424\
 Data File : P0102308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2024 14:56
 Operator : YP/AJ
 Sample : P1669-09MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-6MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:26:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.274	3.388	85726381	57313121	18.485	17.638
2) SA Decachlor...	9.816	8.256	50954743	34585551	20.010	16.718
Target Compounds						
3) L1 AR-1016-1	5.418	4.440	59732275	42121414	462.843	459.024
4) L1 AR-1016-2	5.440	4.458	87332399	59185037	456.558	441.522
5) L1 AR-1016-3	5.501	4.630	55570729	33771286	450.067	457.942
6) L1 AR-1016-4	5.598	4.671	45818481	28285279	467.177	448.794
7) L1 AR-1016-5	5.888	4.879	45853814	31419052	473.145	398.963
31) L7 AR-1260-1	6.999	5.892	73828160	61402983	454.353	414.610
32) L7 AR-1260-2	7.253	6.078	78331721	70864733	439.245	412.826
33) L7 AR-1260-3	7.608	6.228	50070102	63615655	384.771	384.783
34) L7 AR-1260-4	7.832	6.692	60635568	43674971	426.877	342.188
35) L7 AR-1260-5	8.136	6.933	105.5E6	99713788	399.480	364.261



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030424\
Data File : PO102308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 14:56
Operator : YP/AJ
Sample : P1669-09MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-6MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 00:26:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.274 min
Delta R.T.: -0.008 min
Response: 85726381
Conc: 18.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#1 Tetrachloro-m-xylene

R.T.: 3.388 min
Delta R.T.: -0.004 min
Response: 57313121
Conc: 17.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.816 min
Delta R.T.: -0.025 min
Response: 50954743
Conc: 20.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.256 min
Delta R.T.: -0.016 min
Response: 34585551
Conc: 16.72 ng/ml



#3 AR-1016-1


R.T.: 5.418 min
Delta R.T.: -0.011 min
Response: 59732275
Conc: 462.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#3 AR-1016-1

R.T.: 4.440 min
Delta R.T.: -0.007 min
Response: 42121414
Conc: 459.02 ng/ml




#4 AR-1016-2

R.T.: 5.440 min
Delta R.T.: -0.010 min
Response: 87332399
Conc: 456.56 ng/ml



#4 AR-1016-2


R.T.: 4.458 min
Delta R.T.: -0.007 min
Response: 59185037
Conc: 441.52 ng/ml



#5 AR-1016-3


R.T.: 5.501 min
Delta R.T.: -0.010 min
Response: 55570729
Conc: 450.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#5 AR-1016-3

R.T.: 4.630 min
Delta R.T.: -0.007 min
Response: 33771286
Conc: 457.94 ng/ml




#6 AR-1016-4

R.T.: 5.598 min
Delta R.T.: -0.010 min
Response: 45818481
Conc: 467.18 ng/ml



#6 AR-1016-4


R.T.: 4.671 min
Delta R.T.: -0.007 min
Response: 28285279
Conc: 448.79 ng/ml



#7 AR-1016-5


R.T.: 5.888 min
Delta R.T.: -0.012 min
Response: 45853814
Conc: 473.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#7 AR-1016-5

R.T.: 4.879 min
Delta R.T.: -0.008 min
Response: 31419052
Conc: 398.96 ng/ml




#31 AR-1260-1

R.T.: 6.999 min
Delta R.T.: -0.014 min
Response: 73828160
Conc: 454.35 ng/ml



#31 AR-1260-1


R.T.: 5.892 min
Delta R.T.: -0.010 min
Response: 61402983
Conc: 414.61 ng/ml



#32 AR-1260-2


R.T.: 7.253 min
Delta R.T.: -0.013 min
Response: 78331721
Conc: 439.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#32 AR-1260-2

R.T.: 6.078 min
Delta R.T.: -0.010 min
Response: 70864733
Conc: 412.83 ng/ml




#33 AR-1260-3

R.T.: 7.608 min
Delta R.T.: -0.014 min
Response: 50070102
Conc: 384.77 ng/ml



#33 AR-1260-3


R.T.: 6.228 min
Delta R.T.: -0.011 min
Response: 63615655
Conc: 384.78 ng/ml



#34 AR-1260-4


R.T.: 7.832 min
Delta R.T.: -0.015 min
Response: 60635568
Conc: 426.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6MS




#34 AR-1260-4

R.T.: 6.692 min
Delta R.T.: -0.011 min
Response: 43674971
Conc: 342.19 ng/ml



#35 AR-1260-5

R.T.: 8.136 min
Delta R.T.: -0.016 min
Response: 105489299
Conc: 399.48 ng/ml



#35 AR-1260-5

R.T.: 6.933 min
Delta R.T.: -0.012 min
Response: 99713788
Conc: 364.26 ng/ml