

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030624\
 Data File : P0102351.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2024 15:27
 Operator : YP/AJ
 Sample : P1682-01MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HR-04-03052024MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/07/2024
 Supervised By :Ankita Jodhani 03/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 00:32:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.274	3.388	79006195	49699361	17.036	15.295
2)	SA Decachlor...	9.822	8.257	49016494	31243104	19.249	15.102
Target Compounds							
3)	L1 AR-1016-1	5.418	4.440	56014261	30614619	434.033m	333.627m
4)	L1 AR-1016-2	5.440	4.458	76865220	45867274	401.838m	342.171m
5)	L1 AR-1016-3	5.501	4.628	48740384	24646744	394.748m	334.213m
6)	L1 AR-1016-4	5.598	4.671	42325651	21768660	431.563m	345.397m
7)	L1 AR-1016-5	5.889	4.879	37033419	29044960	382.131m	368.817
31)	L7 AR-1260-1	7.000	5.892	74048162	53392881	455.707	360.524
32)	L7 AR-1260-2	7.253	6.078	75594982	64487995	423.899m	375.678
33)	L7 AR-1260-3	7.611	6.227	46933666	58542784	360.669	354.099
34)	L7 AR-1260-4	7.833	6.693	59461665	41079816	418.613	321.856
35)	L7 AR-1260-5	8.138	6.934	100.5E6	91479961	380.677	334.182

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030624\
Data File : PO102351.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2024 15:27
Operator : YP/AJ
Sample : P1682-01MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HR-04-03052024MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/07/2024
Supervised By :Ankita Jodhani 03/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 00:32:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

