

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030624\
 Data File : P0102357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2024 17:38
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 00:38:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.275	3.388	196.1E6	145.2E6	42.295	44.698
2)	SA Decachlor...	9.821	8.258	134.6E6	83819087	52.865	40.516
Target Compounds							
16)	L4 AR-1242-1	5.420	4.441	47900537	34959483	487.516	494.980
17)	L4 AR-1242-2	5.442	4.458	70316718	50005372	484.193	492.556
18)	L4 AR-1242-3	5.503	4.630	43720146	27960452	472.249	497.661
19)	L4 AR-1242-4	5.600	4.712	35509127	28498116	481.154	476.441
20)	L4 AR-1242-5	6.326	5.223	38977885	31279516	556.640	473.878



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030624\
Data File : PO102357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2024 17:38
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 00:38:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.275 min
Delta R.T.: -0.007 min
Response: 196146657
Conc: 42.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.388 min
Delta R.T.: -0.004 min
Response: 145243999
Conc: 44.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.821 min
Delta R.T.: -0.020 min
Response: 134615996
Conc: 52.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.258 min
Delta R.T.: -0.014 min
Response: 83819087
Conc: 40.52 ng/ml



#16 AR-1242-1


R.T.: 5.420 min
Delta R.T.: -0.008 min
Response: 47900537
Conc: 487.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.441 min
Delta R.T.: -0.007 min
Response: 34959483
Conc: 494.98 ng/ml




#17 AR-1242-2

R.T.: 5.442 min
Delta R.T.: -0.009 min
Response: 70316718
Conc: 484.19 ng/ml



#17 AR-1242-2


R.T.: 4.458 min
Delta R.T.: -0.007 min
Response: 50005372
Conc: 492.56 ng/ml



#18 AR-1242-3


R.T.: 5.503 min
Delta R.T.: -0.009 min
Response: 43720146
Conc: 472.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.630 min
Delta R.T.: -0.007 min
Response: 27960452
Conc: 497.66 ng/ml




#19 AR-1242-4

R.T.: 5.600 min
Delta R.T.: -0.008 min
Response: 35509127
Conc: 481.15 ng/ml



#19 AR-1242-4


R.T.: 4.712 min
Delta R.T.: -0.007 min
Response: 28498116
Conc: 476.44 ng/ml



#20 AR-1242-5

R.T.: 6.326 min
Delta R.T.: -0.010 min
Response: 38977885
Conc: 556.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.223 min
Delta R.T.: -0.008 min
Response: 31279516
Conc: 473.88 ng/ml