

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030625\
 Data File : P0109649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2025 10:24
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 10:36:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 04:40:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.697	3.694	467.0E6	269.5E6	49.339	51.479
2) SA Decachlor...	8.757	8.707	362.2E6	141.0E6	42.103	44.268
Target Compounds						
16) L4 AR-1242-1	4.791	4.778	129.6E6	70206143	495.211	530.470
17) L4 AR-1242-2	4.810	4.797	177.7E6	98270714	499.113	541.390
18) L4 AR-1242-3	4.867	4.973	124.8E6	54312221	496.461	543.804
19) L4 AR-1242-4	4.988	5.057	98239539	55771322	500.069	521.625
20) L4 AR-1242-5	5.642	5.579	106.0E6	67585492	494.678	537.650



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030625\
Data File : PO109649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2025 10:24
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 10:36:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.697 min
Delta R.T.: -0.001 min
Response: 466997941
Conc: 49.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 269450006
Conc: 51.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.757 min
Delta R.T.: 0.000 min
Response: 362162701
Conc: 42.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.707 min
Delta R.T.: 0.000 min
Response: 140959437
Conc: 44.27 ng/ml



#16 AR-1242-1


R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 129628779
Conc: 495.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 70206143
Conc: 530.47 ng/ml




#17 AR-1242-2

R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 177707329
Conc: 499.11 ng/ml



#17 AR-1242-2


R.T.: 4.797 min
Delta R.T.: 0.000 min
Response: 98270714
Conc: 541.39 ng/ml



#18 AR-1242-3


R.T.: 4.867 min
Delta R.T.: 0.000 min
Response: 124795463
Conc: 496.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.973 min
Delta R.T.: 0.000 min
Response: 54312221
Conc: 543.80 ng/ml




#19 AR-1242-4

R.T.: 4.988 min
Delta R.T.: -0.001 min
Response: 98239539
Conc: 500.07 ng/ml



#19 AR-1242-4


R.T.: 5.057 min
Delta R.T.: 0.000 min
Response: 55771322
Conc: 521.62 ng/ml



#20 AR-1242-5

R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 106047200
Conc: 494.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.579 min
Delta R.T.: 0.000 min
Response: 67585492
Conc: 537.65 ng/ml