

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030723\
 Data File : P0093030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2023 16:38
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/08/2023
 Supervised By :Ankita Jodhani 03/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 17:24:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.635	159.9E6	60506668	53.266	47.468
2) SA Decachlor...	10.266	8.669	112.5E6	53854168	58.500	46.211
Target Compounds						
26) L6 AR-1254-1	6.470	5.525	50340128	35177100	443.989m	502.450m
27) L6 AR-1254-2	6.690	5.673	84980190	29552111	511.299m	471.035m
28) L6 AR-1254-3	7.059	6.078	84889565	43342519	518.565m	449.624m
29) L6 AR-1254-4	7.347	6.309	50942218	25170652	490.853m	487.781m
30) L6 AR-1254-5	7.771	6.728	62353260	40173613	508.576m	494.803

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030723\
Data File : PO093030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2023 16:38
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/08/2023
Supervised By :Ankita Jodhani 03/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 17:24:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.430 min
Delta R.T.: -0.004 min
Response: 159944508
Conc: 53.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/08/2023
Supervised By :Ankita Jodhani 03/08/2023




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: -0.003 min
Response: 60506668
Conc: 47.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.266 min
Delta R.T.: 0.009 min
Response: 112523847
Conc: 58.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.669 min
Delta R.T.: -0.003 min
Response: 53854168
Conc: 46.21 ng/ml




#26 AR-1254-1

R.T.: 6.470 min
Delta R.T.: 0.000 min
Response: 50340128
Conc: 443.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/08/2023
Supervised By :Ankita Jodhani 03/08/2023




#26 AR-1254-1

R.T.: 5.525 min
Delta R.T.: -0.003 min
Response: 35177100
Conc: 502.45 ng/ml m




#27 AR-1254-2

R.T.: 6.690 min
Delta R.T.: -0.002 min
Response: 84980190
Conc: 511.30 ng/ml m



#27 AR-1254-2

R.T.: 5.673 min
Delta R.T.: -0.003 min
Response: 29552111
Conc: 471.03 ng/ml m




#28 AR-1254-3

R.T.: 7.059 min
Delta R.T.: -0.002 min
Response: 84889565
Conc: 518.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/08/2023
Supervised By :Ankita Jodhani 03/08/2023




#28 AR-1254-3

R.T.: 6.078 min
Delta R.T.: -0.004 min
Response: 43342519
Conc: 449.62 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: -0.001 min
Response: 50942218
Conc: 490.85 ng/ml



#29 AR-1254-4

R.T.: 6.309 min
Delta R.T.: -0.002 min
Response: 25170652
Conc: 487.78 ng/ml




#30 AR-1254-5

R.T.: 7.771 min
Delta R.T.: 0.000 min
Response: 62353260
Conc: 508.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/08/2023
Supervised By :Ankita Jodhani 03/08/2023



#30 AR-1254-5

R.T.: 6.728 min
Delta R.T.: -0.003 min
Response: 40173613
Conc: 494.80 ng/ml