

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030724\
 Data File : P0102393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2024 19:18
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 21:58:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 25 11:24:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.273	3.387	208.2E6	141.5E6	44.898	43.541
2)	SA Decachlor...	9.817	8.256	132.4E6	86963568	52.014	42.036
Target Compounds							
26)	L6 AR-1254-1	6.260	5.222	69674170	65637287	518.707	451.057
27)	L6 AR-1254-2	6.475	5.367	100.4E6	56551404	509.325	438.359
28)	L6 AR-1254-3	6.837	5.764	101.0E6	86734520	520.691	444.098
29)	L6 AR-1254-4	7.119	5.990	67908995	52258465	548.199	485.329
30)	L6 AR-1254-5	7.535	6.400	72110347	73850244	513.617	451.734



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030724\
Data File : PO102393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2024 19:18
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 21:58:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 25 11:24:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.273 min
Delta R.T.: -0.009 min
Response: 208216649
Conc: 44.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.387 min
Delta R.T.: -0.005 min
Response: 141484225
Conc: 43.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.817 min
Delta R.T.: -0.024 min
Response: 132448595
Conc: 52.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.256 min
Delta R.T.: -0.016 min
Response: 86963568
Conc: 42.04 ng/ml



#26 AR-1254-1


R.T.: 6.260 min
Delta R.T.: -0.012 min
Response: 69674170
Conc: 518.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.222 min
Delta R.T.: -0.010 min
Response: 65637287
Conc: 451.06 ng/ml




#27 AR-1254-2

R.T.: 6.475 min
Delta R.T.: -0.012 min
Response: 100407422
Conc: 509.32 ng/ml



#27 AR-1254-2


R.T.: 5.367 min
Delta R.T.: -0.010 min
Response: 56551404
Conc: 438.36 ng/ml



#28 AR-1254-3


R.T.: 6.837 min
Delta R.T.: -0.013 min
Response: 100962810
Conc: 520.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.764 min
Delta R.T.: -0.011 min
Response: 86734520
Conc: 444.10 ng/ml




#29 AR-1254-4

R.T.: 7.119 min
Delta R.T.: -0.015 min
Response: 67908995
Conc: 548.20 ng/ml



#29 AR-1254-4


R.T.: 5.990 min
Delta R.T.: -0.011 min
Response: 52258465
Conc: 485.33 ng/ml



#30 AR-1254-5

R.T.: 7.535 min
Delta R.T.: -0.013 min
Response: 72110347
Conc: 513.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.400 min
Delta R.T.: -0.012 min
Response: 73850244
Conc: 451.73 ng/ml