

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030821\
 Data File : P0075955.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Mar 2021 16:23
 Operator : DD\AJ
 Sample : M1581-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 03-05E

Manual Integrations
 APPROVED

Ankita
 3/9/2021 1:04:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 07:16:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 08:07:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.924	3.941	1469135	1083972	24.647m	24.300
2) SA Decachlor...	10.928	9.213	897046	900228	27.306	24.758

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030821\
Data File : P0075955.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Mar 2021 16:23
Operator : DD\AJ
Sample : M1581-01
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
03-05E


Manual Integrations
APPROVED

Ankita
3/9/2021 1:04:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 07:16:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 08:07:39 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.924 min
Delta R.T.: -0.004 min
Response: 1469135
Conc: 24.65 ng/ml m

Instrument :
ECD_O
ClientSampleId :
03-05E


Manual Integrations
APPROVED

Ankita
3/9/2021 1:04:26 PM




#1 Tetrachloro-m-xylene

R.T.: 3.941 min
Delta R.T.: -0.002 min
Response: 1083972
Conc: 24.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.928 min
Delta R.T.: -0.002 min
Response: 897046
Conc: 27.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.213 min
Delta R.T.: -0.010 min
Response: 900228
Conc: 24.76 ng/ml