

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0030822.M
 Title : GC EXTRACTABLES
 Last Update : Wed Mar 09 03:28:38 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0085149.D 1000 =P0085147.D 750 =P0085148.D
 250 =P0085150.D 50 =P0085151.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.567	1.482	1.506	1.589	1.552	1.539 E5	2.88
2) SA	Decachlorobiphenyl	6.466	6.326	6.446	6.487	5.446	6.234 E4	7.14
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.480	3.243	3.315	3.606	3.452	3.419 E3	4.17
27) L6	AR-1254-2	5.166	4.836	4.960	5.332	5.116	5.082 E3	3.76
28) L6	AR-1254-3	5.185	4.951	5.042	5.243	4.761	5.036 E3	3.82
29) L6	AR-1254-4	3.738	3.568	3.645	3.820	3.517	3.658 E3	3.37
30) L6	AR-1254-5	4.026	3.881	3.951	4.047	3.665	3.914 E3	3.93
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0085149.D 1000 =P0085147.D 750 =P0085148.D
 250 =P0085150.D 50 =P0085151.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.864	4.611	4.689	4.904	4.726	4.759 E4	2.57
2) SA	Decachlorobiphenyl	3.908	3.594	3.709	4.135	4.122	3.893 E4	6.22
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0030822.M
 Title : GC EXTRACTABLES
 Last Update : Wed Mar 09 03:28:38 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0085149.D 1000 =P0085147.D 750 =P0085148.D
 250 =P0085150.D 50 =P0085151.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.526	2.309	2.393	2.703	2.818	2.550 E3	8.28
27) L6	AR-1254-2	2.298	2.074	2.160	2.475	2.671	2.336 E3	10.32
28) L6	AR-1254-3	3.599	3.330	3.433	3.792	3.919	3.615 E3	6.75
29) L6	AR-1254-4	2.227	2.050	2.121	2.358	2.391	2.229 E3	6.59
30) L6	AR-1254-5	3.386	3.128	3.232	3.501	3.551	3.360 E3	5.32
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range