

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030922\
 Data File : PR053872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2022 13:03
 Operator : AJ\MA
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 04:38:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 16 05:21:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.563	2.811	40439174	26997477	20.644	22.451
2)	SA Decachlor...	9.371	7.953	73428710	55644667	45.074	41.473
Target Compounds							
3)	L1 AR-1016-1	4.781	3.893	15978542	11039293	248.620	245.702
4)	L1 AR-1016-2	4.804	3.910	19734968	12386390	216.659	187.001
5)	L1 AR-1016-3	4.868	4.084	9682429	7942940	171.572	223.926 #
6)	L1 AR-1016-4	4.970	4.134	10473043	17767384	226.442	563.523 #
7)	L1 AR-1016-5	5.281	4.345	25171573	20549843	531.924	544.672
9)	L2 AR-1221-2	3.876	3.186f	747764	902612	44.333	83.520 #
10)	L2 AR-1221-3	3.957	3.186	1608487	902612	29.349	24.424
11)	L3 AR-1232-1	3.957	3.186	1608487	902612	35.126	30.165
12)	L3 AR-1232-2	4.500	3.910	7268786	12386390	321.408	428.947 #
13)	L3 AR-1232-3	4.804	4.084	19734968	7942940	488.397	522.259
14)	L3 AR-1232-4	4.970	4.174	10473043	17054282	510.364	1217.104 #
15)	L3 AR-1232-5	5.068	4.345	22615340	20549843	1445.291	1294.992
16)	L4 AR-1242-1	4.781	3.893	15978542	11039293	299.157	303.299
17)	L4 AR-1242-2	4.804	3.910	19734968	12386390	263.947	231.716
18)	L4 AR-1242-3	4.868	4.084	9682429	7942940	205.966	276.038 #
19)	L4 AR-1242-4	4.970	4.174	10473043	17054282	271.142	592.377 #
20)	L4 AR-1242-5	5.751	4.707	29628339	28164662	697.023	781.346
21)	L5 AR-1248-1	4.781	3.893	15978542	11039293	399.412	406.990
22)	L5 AR-1248-2	5.068	4.134	22615340	17767384	407.725	413.096
23)	L5 AR-1248-3	5.281	4.174	25171573	17054282	386.910	406.987
24)	L5 AR-1248-4	5.711	4.345	29467099	20549843	400.128	404.692
25)	L5 AR-1248-5	5.751	4.748	29628339	19812957	410.450	401.754
26)	L6 AR-1254-1	5.679	4.707	23095939	28164662	323.283	375.446
27)	L6 AR-1254-2	5.917	4.863	29618296	8974437	262.931	136.523 #
28)	L6 AR-1254-3	6.307	5.274	16764809	13078293	142.793	126.843
29)	L6 AR-1254-4	6.617	5.518	11477165	8835060	126.455	134.215
30)	L6 AR-1254-5	7.068	5.954	3097754	2669185	31.954	28.363
31)	L7 AR-1260-1	6.480	5.417	1368735	6367766	16.211	89.598 #
32)	L7 AR-1260-2	6.759	5.615	2098173	884548	21.364	10.848 #
33)	L7 AR-1260-3	7.133	5.763	529534	1472490	7.077	18.745 #
34)	L7 AR-1260-4	7.382	6.299	1799578	797866	20.265	11.860 #
35)	L7 AR-1260-5	7.730	0.000	923136	0	5.396	N.D. #
36)	L8 AR-1262-1	7.133	5.954	529534	2669185	4.632	51.545 #
37)	L8 AR-1262-2	7.730	0.000	923136	0	4.491	N.D. #
38)	L8 AR-1262-3	8.042	0.000	1306065	0	13.812	N.D. #
39)	L8 AR-1262-4	8.100	0.000	745713	0	13.383	N.D. #
41)	L9 AR-1268-1	8.042	0.000	1306065	0	5.352	N.D. #
42)	L9 AR-1268-2	8.100	0.000	745713	0	3.326	N.D. #
45)	L9 AR-1268-5	9.085	0.000	307174	0	0.495	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030922\
Data File : PR053872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2022 13:03
Operator : AJ\MA
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 04:38:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR021522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 16 05:21:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

