

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030823\
 Data File : P0093042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2023 11:02
 Operator : YP/AJ
 Sample : 01827-01MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EO-03-030623MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/09/2023
 Supervised By :Ankita Jodhani 03/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 16:15:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.429	3.634	62140374	23687166	20.694	18.583m
2) SA Decachlor...	10.261	8.666	46707583	18852660	24.283	16.177 #
Target Compounds						
3) L1 AR-1016-1	5.605	4.720	43771315	17025837	476.248	421.450
4) L1 AR-1016-2	5.628	4.738	61943736	24285668	464.760	420.590
5) L1 AR-1016-3	5.690	4.914	38757892	13271674	446.407	423.526
6) L1 AR-1016-4	5.789	4.957	30324139	11621961	464.365	415.707
7) L1 AR-1016-5	6.087	5.170	33106634	14504001	465.831	408.877
31) L7 AR-1260-1	7.224	6.208	57522490	26719050	471.381	392.187
32) L7 AR-1260-2	7.482	6.396	68155062	31544710	463.705	392.432
33) L7 AR-1260-3	7.767	6.551	81273533	28136911	517.051m	348.571 #
34) L7 AR-1260-4	8.073	7.025	54783319	20278056	467.689	333.129 #
35) L7 AR-1260-5	8.400	7.268	97470558	45164348	469.990	353.761

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030823\
Data File : PO093042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2023 11:02
Operator : YP/AJ
Sample : 01827-01MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EO-03-030623MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/09/2023
Supervised By :Ankita Jodhani 03/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 16:15:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

