

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0030922\
 Data File : P0085211.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Mar 2022 22:15
 Operator : YP\AJ
 Sample : N1645-01
 Misc : AR1660 LOD 40 PPB
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:22:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0030822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 09 04:29:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.499	3.591	4054528	1306556	23.614	28.633
2)	SA Decachlor...	10.415	8.610	1884540	1184406	28.779	31.110
Target Compounds							
3)	L1 AR-1016-1	5.688	4.680	145136	59852	31.822	35.748
4)	L1 AR-1016-2	5.711	4.698	181268	89730	26.911	37.055 #
5)	L1 AR-1016-3	5.775	4.874	115421	49685	28.586	36.295 #
6)	L1 AR-1016-4	5.875	4.918	82082	41095	24.500	35.784 #
7)	L1 AR-1016-5	6.174	5.130	92331	50233	29.918	35.074
8)	L2 AR-1221-1	4.710	3.804	103566	6126	56.048	11.113 #
9)	L2 AR-1221-2	4.810	3.890	83806	32355	63.002	76.101
10)	L2 AR-1221-3	4.877	3.968	117339	30668	30.079	23.544
11)	L3 AR-1232-1	4.877	3.968	117339	30668	41.560	30.877 #
12)	L3 AR-1232-2	5.415	4.698	132732	89730	82.393	85.725
13)	L3 AR-1232-3	5.711	4.874	181268	49685	64.038	87.265 #
14)	L3 AR-1232-4	5.875	4.959	82082	48521	60.188	91.876 #
15)	L3 AR-1232-5	5.967	5.130	65621	50233	63.789	85.021 #
16)	L4 AR-1242-1	5.688	4.680	145136	59852	42.592	47.808
17)	L4 AR-1242-2	5.711	4.698	181268	89730	36.209	49.958 #
18)	L4 AR-1242-3	5.775	4.874	115421	49685	38.010	50.352 #
19)	L4 AR-1242-4	5.875	4.959	82082	48521	34.841	47.449 #
20)	L4 AR-1242-5	6.627	5.483	6089	38594	3.218	30.057 #
21)	L5 AR-1248-1	5.688	4.680	145136	59852	54.614	58.639
22)	L5 AR-1248-2	5.967	4.918	65621	41095	19.778	28.365 #
23)	L5 AR-1248-3	6.174	4.959	92331	48521	26.943	32.804
24)	L5 AR-1248-4	6.584	5.130	9312	50233	2.807	28.776 #
25)	L5 AR-1248-5	6.627	5.483f	6089	38594	1.926	23.581 #
26)	L6 AR-1254-1	6.556	5.483	55453	38594	16.218	15.135
27)	L6 AR-1254-2	6.780	5.632	70626	36698	13.898	15.712
28)	L6 AR-1254-3	7.150	6.052	57911	75778	11.498	20.964 #
29)	L6 AR-1254-4	7.441	6.265	65427	11433	17.888	5.128 #
30)	L6 AR-1254-5	7.864	6.683	214413	117944	54.779	35.106 #
31)	L7 AR-1260-1	7.318	6.166	156313	95517	38.426	36.385
32)	L7 AR-1260-2	7.578	6.356	188055	114927	39.231	36.754
33)	L7 AR-1260-3	7.941	6.508	146135	108663	44.967	35.085
34)	L7 AR-1260-4	8.174	6.983	164808	84116	43.120	38.180
35)	L7 AR-1260-5	8.505	7.226	253138	191293	36.772	38.420
36)	L8 AR-1262-1	7.941	6.683	146135	117944	32.074	68.658 #
37)	L8 AR-1262-2	8.505	6.983	253138	84116	30.839	29.208
38)	L8 AR-1262-3	8.845	7.511	131202	45819	24.527	20.278
39)	L8 AR-1262-4	8.928	7.574	34491	133933	13.867	32.557 #
40)	L8 AR-1262-5	9.618	8.071	60016	48903	21.690	26.258
41)	L9 AR-1268-1	8.845	7.511	131202	45819	13.997	6.985 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030922\
Data File : P0085211.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Mar 2022 22:15
Operator : YP\AJ
Sample : N1645-01
Misc : AR1660 LOD 40 PPB
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:22:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 09 04:29:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.928	7.574	34491	133933	4.064	22.897 #
43)	L9 AR-1268-3	9.171	7.784	7598	5165	1.059	1.042
44)	L9 AR-1268-4	9.618	8.071	60016	48903	19.607	23.044
45)	L9 AR-1268-5	10.050	8.359	13398	20281	0.562	1.413 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030922\
Data File : P0085211.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Mar 2022 22:15
Operator : YP\AJ
Sample : N1645-01
Misc : AR1660 LOD 40 PPB
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:22:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 09 04:29:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.499 min
Delta R.T.: -0.001 min
Response: 4054528
Conc: 23.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.001 min
Response: 1306556
Conc: 28.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.415 min
Delta R.T.: -0.003 min
Response: 1884540
Conc: 28.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.610 min
Delta R.T.: -0.003 min
Response: 1184406
Conc: 31.11 ng/ml



#3 AR-1016-1


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 145136
Conc: 31.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.680 min
Delta R.T.: -0.001 min
Response: 59852
Conc: 35.75 ng/ml




#4 AR-1016-2

R.T.: 5.711 min
Delta R.T.: -0.002 min
Response: 181268
Conc: 26.91 ng/ml



#4 AR-1016-2


R.T.: 4.698 min
Delta R.T.: -0.002 min
Response: 89730
Conc: 37.06 ng/ml



#5 AR-1016-3


R.T.: 5.775 min
Delta R.T.: -0.001 min
Response: 115421
Conc: 28.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.874 min
Delta R.T.: -0.002 min
Response: 49685
Conc: 36.29 ng/ml




#6 AR-1016-4

R.T.: 5.875 min
Delta R.T.: -0.001 min
Response: 82082
Conc: 24.50 ng/ml



#6 AR-1016-4


R.T.: 4.918 min
Delta R.T.: -0.002 min
Response: 41095
Conc: 35.78 ng/ml



#7 AR-1016-5


R.T.: 6.174 min
Delta R.T.: -0.002 min
Response: 92331
Conc: 29.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.130 min
Delta R.T.: -0.002 min
Response: 50233
Conc: 35.07 ng/ml




#8 AR-1221-1

R.T.: 4.710 min
Delta R.T.: 0.000 min
Response: 103566
Conc: 56.05 ng/ml



#8 AR-1221-1


R.T.: 3.804 min
Delta R.T.: -0.003 min
Response: 6126
Conc: 11.11 ng/ml



#9 AR-1221-2


R.T.: 4.810 min
Delta R.T.: 0.011 min
Response: 83806
Conc: 63.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.890 min
Delta R.T.: -0.002 min
Response: 32355
Conc: 76.10 ng/ml




#10 AR-1221-3

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 117339
Conc: 30.08 ng/ml



#10 AR-1221-3


R.T.: 3.968 min
Delta R.T.: 0.000 min
Response: 30668
Conc: 23.54 ng/ml



#11 AR-1232-1


R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 117339
Conc: 41.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.968 min
Delta R.T.: 0.000 min
Response: 30668
Conc: 30.88 ng/ml




#12 AR-1232-2

R.T.: 5.415 min
Delta R.T.: -0.001 min
Response: 132732
Conc: 82.39 ng/ml



#12 AR-1232-2


R.T.: 4.698 min
Delta R.T.: -0.001 min
Response: 89730
Conc: 85.72 ng/ml



#13 AR-1232-3


R.T.: 5.711 min
Delta R.T.: 0.000 min
Response: 181268
Conc: 64.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.874 min
Delta R.T.: 0.000 min
Response: 49685
Conc: 87.27 ng/ml




#14 AR-1232-4

R.T.: 5.875 min
Delta R.T.: 0.000 min
Response: 82082
Conc: 60.19 ng/ml



#14 AR-1232-4


R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 48521
Conc: 91.88 ng/ml



#15 AR-1232-5


R.T.: 5.967 min
Delta R.T.: 0.000 min
Response: 65621
Conc: 63.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.130 min
Delta R.T.: -0.001 min
Response: 50233
Conc: 85.02 ng/ml




#16 AR-1242-1

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 145136
Conc: 42.59 ng/ml



#16 AR-1242-1


R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 59852
Conc: 47.81 ng/ml



#17 AR-1242-2


R.T.: 5.711 min
Delta R.T.: 0.000 min
Response: 181268
Conc: 36.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.698 min
Delta R.T.: 0.000 min
Response: 89730
Conc: 49.96 ng/ml




#18 AR-1242-3

R.T.: 5.775 min
Delta R.T.: 0.000 min
Response: 115421
Conc: 38.01 ng/ml



#18 AR-1242-3


R.T.: 4.874 min
Delta R.T.: 0.000 min
Response: 49685
Conc: 50.35 ng/ml



#19 AR-1242-4


R.T.: 5.875 min
Delta R.T.: 0.000 min
Response: 82082
Conc: 34.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 48521
Conc: 47.45 ng/ml




#20 AR-1242-5

R.T.: 6.627 min
Delta R.T.: 0.004 min
Response: 6089
Conc: 3.22 ng/ml



#20 AR-1242-5


R.T.: 5.483 min
Delta R.T.: 0.000 min
Response: 38594
Conc: 30.06 ng/ml



#21 AR-1248-1


R.T.: 5.688 min
Delta R.T.: -0.001 min
Response: 145136
Conc: 54.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 59852
Conc: 58.64 ng/ml




#22 AR-1248-2

R.T.: 5.967 min
Delta R.T.: 0.000 min
Response: 65621
Conc: 19.78 ng/ml



#22 AR-1248-2


R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 41095
Conc: 28.36 ng/ml



#23 AR-1248-3


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 92331
Conc: 26.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 48521
Conc: 32.80 ng/ml




#24 AR-1248-4

R.T.: 6.584 min
Delta R.T.: 0.000 min
Response: 9312
Conc: 2.81 ng/ml



#24 AR-1248-4


R.T.: 5.130 min
Delta R.T.: 0.000 min
Response: 50233
Conc: 28.78 ng/ml



#25 AR-1248-5


R.T.: 6.627 min
Delta R.T.: 0.004 min
Response: 6089
Conc: 1.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.483 min
Delta R.T.: -0.041 min
Response: 38594
Conc: 23.58 ng/ml




#26 AR-1254-1

R.T.: 6.556 min
Delta R.T.: -0.003 min
Response: 55453
Conc: 16.22 ng/ml



#26 AR-1254-1


R.T.: 5.483 min
Delta R.T.: 0.000 min
Response: 38594
Conc: 15.14 ng/ml



#27 AR-1254-2


R.T.: 6.780 min
Delta R.T.: 0.000 min
Response: 70626
Conc: 13.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.632 min
Delta R.T.: 0.000 min
Response: 36698
Conc: 15.71 ng/ml




#28 AR-1254-3

R.T.: 7.150 min
Delta R.T.: -0.002 min
Response: 57911
Conc: 11.50 ng/ml



#28 AR-1254-3


R.T.: 6.052 min
Delta R.T.: 0.015 min
Response: 75778
Conc: 20.96 ng/ml



#29 AR-1254-4


R.T.: 7.441 min
Delta R.T.: 0.000 min
Response: 65427
Conc: 17.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.265 min
Delta R.T.: 0.000 min
Response: 11433
Conc: 5.13 ng/ml




#30 AR-1254-5

R.T.: 7.864 min
Delta R.T.: -0.002 min
Response: 214413
Conc: 54.78 ng/ml



#30 AR-1254-5


R.T.: 6.683 min
Delta R.T.: -0.002 min
Response: 117944
Conc: 35.11 ng/ml



#31 AR-1260-1


R.T.: 7.318 min
Delta R.T.: -0.001 min
Response: 156313
Conc: 38.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.166 min
Delta R.T.: -0.003 min
Response: 95517
Conc: 36.38 ng/ml




#32 AR-1260-2

R.T.: 7.578 min
Delta R.T.: 0.000 min
Response: 188055
Conc: 39.23 ng/ml



#32 AR-1260-2


R.T.: 6.356 min
Delta R.T.: -0.003 min
Response: 114927
Conc: 36.75 ng/ml



#33 AR-1260-3


R.T.: 7.941 min
Delta R.T.: -0.001 min
Response: 146135
Conc: 44.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.508 min
Delta R.T.: -0.003 min
Response: 108663
Conc: 35.09 ng/ml




#34 AR-1260-4

R.T.: 8.174 min
Delta R.T.: 0.000 min
Response: 164808
Conc: 43.12 ng/ml



#34 AR-1260-4


R.T.: 6.983 min
Delta R.T.: -0.002 min
Response: 84116
Conc: 38.18 ng/ml



#35 AR-1260-5


R.T.: 8.505 min
Delta R.T.: -0.001 min
Response: 253138
Conc: 36.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.226 min
Delta R.T.: -0.003 min
Response: 191293
Conc: 38.42 ng/ml




#36 AR-1262-1

R.T.: 7.941 min
Delta R.T.: 0.000 min
Response: 146135
Conc: 32.07 ng/ml



#36 AR-1262-1


R.T.: 6.683 min
Delta R.T.: 0.000 min
Response: 117944
Conc: 68.66 ng/ml



#37 AR-1262-2


R.T.: 8.505 min
Delta R.T.: 0.002 min
Response: 253138
Conc: 30.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 84116
Conc: 29.21 ng/ml




#38 AR-1262-3

R.T.: 8.845 min
Delta R.T.: 0.013 min
Response: 131202
Conc: 24.53 ng/ml



#38 AR-1262-3


R.T.: 7.511 min
Delta R.T.: 0.000 min
Response: 45819
Conc: 20.28 ng/ml



#39 AR-1262-4


R.T.: 8.928 min
Delta R.T.: 0.002 min
Response: 34491
Conc: 13.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.574 min
Delta R.T.: -0.002 min
Response: 133933
Conc: 32.56 ng/ml




#40 AR-1262-5

R.T.: 9.618 min
Delta R.T.: 0.004 min
Response: 60016
Conc: 21.69 ng/ml



#40 AR-1262-5


R.T.: 8.071 min
Delta R.T.: -0.001 min
Response: 48903
Conc: 26.26 ng/ml



#41 AR-1268-1


R.T.: 8.845 min
Delta R.T.: 0.017 min
Response: 131202
Conc: 14.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.511 min
Delta R.T.: -0.002 min
Response: 45819
Conc: 6.98 ng/ml




#42 AR-1268-2

R.T.: 8.928 min
Delta R.T.: 0.000 min
Response: 34491
Conc: 4.06 ng/ml



#42 AR-1268-2


R.T.: 7.574 min
Delta R.T.: -0.004 min
Response: 133933
Conc: 22.90 ng/ml



#43 AR-1268-3


R.T.: 9.171 min
Delta R.T.: 0.002 min
Response: 7598
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.784 min
Delta R.T.: -0.001 min
Response: 5165
Conc: 1.04 ng/ml




#44 AR-1268-4

R.T.: 9.618 min
Delta R.T.: 0.003 min
Response: 60016
Conc: 19.61 ng/ml



#44 AR-1268-4


R.T.: 8.071 min
Delta R.T.: -0.002 min
Response: 48903
Conc: 23.04 ng/ml



#45 AR-1268-5

R.T.: 10.050 min
Delta R.T.: -0.004 min
Response: 13398
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.359 min
Delta R.T.: -0.001 min
Response: 20281
Conc: 1.41 ng/ml