

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0030922\
 Data File : P0085215.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Mar 2022 23:19
 Operator : YP\AJ
 Sample : N1645-01
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:25:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0030822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 09 04:29:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.499	3.591	3731333	1200397	21.731	26.306
2)	SA Decachlor...	10.414	8.610	1753922	1104818	26.784	29.019
Target Compounds							
3)	L1 AR-1016-1	5.688	4.680	71575	28564	15.694	17.060
4)	L1 AR-1016-2	5.710	4.697	82211	44025	12.205	18.181 #
5)	L1 AR-1016-3	5.769	4.875	70306	21846	17.412	15.958
6)	L1 AR-1016-4	5.874	4.918	43575	59076	13.006	51.440 #
7)	L1 AR-1016-5	6.175	5.130	102335	60738	33.159	42.410 #
8)	L2 AR-1221-1	4.711	0.000	109326	0	59.165	N.D. #
9)	L2 AR-1221-2	4.811	3.889	89293	27699	67.128	65.150
10)	L2 AR-1221-3	4.871	0.000	53266	0	13.654	N.D. #
11)	L3 AR-1232-1	4.871	0.000	53266	0	18.866	N.D. #
12)	L3 AR-1232-2	5.413	4.697	71077	44025	44.121	42.060
13)	L3 AR-1232-3	5.710	4.875	82211	21846	29.043	38.370 #
14)	L3 AR-1232-4	5.874	4.959	43575	52419	31.952	99.258 #
15)	L3 AR-1232-5	5.968	5.130	86446	60738	84.033	102.801
16)	L4 AR-1242-1	5.688	4.680	71575	28564	21.005	22.816
17)	L4 AR-1242-2	5.710	4.697	82211	44025	16.422	24.511 #
18)	L4 AR-1242-3	5.769	4.875	70306	21846	23.153	22.139
19)	L4 AR-1242-4	5.874	4.959	43575	52419	18.496	51.261 #
20)	L4 AR-1242-5	6.624	5.483	87570	83324	46.284	64.892 #
21)	L5 AR-1248-1	5.688	4.680	71575	28564	26.934	27.985
22)	L5 AR-1248-2	5.968	4.918	86446	59076	26.054	40.776 #
23)	L5 AR-1248-3	6.175	4.959	102335	52419	29.862	35.440
24)	L5 AR-1248-4	6.585	5.130	105719	60738	31.867	34.794
25)	L5 AR-1248-5	6.624	5.524	87570	55647	27.702	34.001
26)	L6 AR-1254-1	6.557	5.483	91548	83324	26.775	32.677
27)	L6 AR-1254-2	6.783	5.632	132104	24702	25.995	10.576 #
28)	L6 AR-1254-3	7.151	6.036	113577	36759	22.551	10.169 #
29)	L6 AR-1254-4	7.442	6.265	151083	45903	41.307	20.589 #
30)	L6 AR-1254-5	7.866	6.686	99023	8714	25.298	2.594 #
31)	L7 AR-1260-1	7.321	6.179	22659	36247	5.570	13.807 #
32)	L7 AR-1260-2	7.596	6.358	83374	6230	17.393	1.992 #
33)	L7 AR-1260-3	7.941	6.497	10863	18178	3.343	5.869 #
34)	L7 AR-1260-4	8.176	7.021f	11495	7443	3.007	3.378
35)	L7 AR-1260-5	8.504	7.228	15049	10011	2.186	2.011
36)	L8 AR-1262-1	7.941	6.686	10863	8714	2.384	5.073 #
37)	L8 AR-1262-2	8.504	7.021f	15049	7443	1.833	2.584 #
38)	L8 AR-1262-3	8.841	7.542	5621	9206	1.051	4.074 #
39)	L8 AR-1262-4	8.941	7.574	6253	1073	2.514	0.261 #
40)	L8 AR-1262-5	9.613	0.000	2116	0	0.765	N.D. #
41)	L9 AR-1268-1	8.841	7.542	5621	9206	0.600	1.403 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030922\
 Data File : P0085215.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Mar 2022 23:19
 Operator : YP\AJ
 Sample : N1645-01
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:25:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 09 04:29:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.941	7.574	6253	1073	0.737	0.183 #
43) L9	AR-1268-3	9.168	0.000	6125	0	0.854	N.D. #
44) L9	AR-1268-4	9.613	0.000	2116	0	0.691	N.D. #
45) L9	AR-1268-5	10.059	8.359	4533	6340	0.190	0.442 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030922\
Data File : P0085215.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Mar 2022 23:19
Operator : YP\AJ
Sample : N1645-01
Misc : AR1248 LOD 40 PPB
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:25:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 09 04:29:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.499 min
Delta R.T.: 0.000 min
Response: 3731333
Conc: 21.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.001 min
Response: 1200397
Conc: 26.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.414 min
Delta R.T.: -0.004 min
Response: 1753922
Conc: 26.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.610 min
Delta R.T.: -0.002 min
Response: 1104818
Conc: 29.02 ng/ml



#3 AR-1016-1


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 71575
Conc: 15.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.680 min
Delta R.T.: -0.002 min
Response: 28564
Conc: 17.06 ng/ml




#4 AR-1016-2

R.T.: 5.710 min
Delta R.T.: -0.002 min
Response: 82211
Conc: 12.20 ng/ml



#4 AR-1016-2


R.T.: 4.697 min
Delta R.T.: -0.003 min
Response: 44025
Conc: 18.18 ng/ml



#5 AR-1016-3


R.T.: 5.769 min
Delta R.T.: -0.007 min
Response: 70306
Conc: 17.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 21846
Conc: 15.96 ng/ml




#6 AR-1016-4

R.T.: 5.874 min
Delta R.T.: -0.002 min
Response: 43575
Conc: 13.01 ng/ml



#6 AR-1016-4


R.T.: 4.918 min
Delta R.T.: -0.002 min
Response: 59076
Conc: 51.44 ng/ml



#7 AR-1016-5


R.T.: 6.175 min
Delta R.T.: -0.001 min
Response: 102335
Conc: 33.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.130 min
Delta R.T.: -0.002 min
Response: 60738
Conc: 42.41 ng/ml




#8 AR-1221-1

R.T.: 4.711 min
Delta R.T.: 0.000 min
Response: 109326
Conc: 59.17 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.807 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.811 min
Delta R.T.: 0.013 min
Response: 89293
Conc: 67.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.889 min
Delta R.T.: -0.003 min
Response: 27699
Conc: 65.15 ng/ml




#10 AR-1221-3

R.T.: 4.871 min
Delta R.T.: -0.005 min
Response: 53266
Conc: 13.65 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.968 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.871 min
Delta R.T.: -0.006 min
Response: 53266
Conc: 18.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.969 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.413 min
Delta R.T.: -0.003 min
Response: 71077
Conc: 44.12 ng/ml



#12 AR-1232-2


R.T.: 4.697 min
Delta R.T.: -0.003 min
Response: 44025
Conc: 42.06 ng/ml



#13 AR-1232-3


R.T.: 5.710 min
Delta R.T.: -0.002 min
Response: 82211
Conc: 29.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 21846
Conc: 38.37 ng/ml




#14 AR-1232-4

R.T.: 5.874 min
Delta R.T.: -0.001 min
Response: 43575
Conc: 31.95 ng/ml



#14 AR-1232-4


R.T.: 4.959 min
Delta R.T.: -0.001 min
Response: 52419
Conc: 99.26 ng/ml



#15 AR-1232-5


R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 86446
Conc: 84.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.130 min
Delta R.T.: -0.001 min
Response: 60738
Conc: 102.80 ng/ml




#16 AR-1242-1

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 71575
Conc: 21.00 ng/ml



#16 AR-1242-1


R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 28564
Conc: 22.82 ng/ml



#17 AR-1242-2


R.T.: 5.710 min
Delta R.T.: 0.000 min
Response: 82211
Conc: 16.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.697 min
Delta R.T.: -0.002 min
Response: 44025
Conc: 24.51 ng/ml




#18 AR-1242-3

R.T.: 5.769 min
Delta R.T.: -0.006 min
Response: 70306
Conc: 23.15 ng/ml



#18 AR-1242-3


R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 21846
Conc: 22.14 ng/ml



#19 AR-1242-4


R.T.: 5.874 min
Delta R.T.: 0.000 min
Response: 43575
Conc: 18.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 52419
Conc: 51.26 ng/ml




#20 AR-1242-5

R.T.: 6.624 min
Delta R.T.: 0.000 min
Response: 87570
Conc: 46.28 ng/ml



#20 AR-1242-5


R.T.: 5.483 min
Delta R.T.: 0.000 min
Response: 83324
Conc: 64.89 ng/ml



#21 AR-1248-1


R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 71575
Conc: 26.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 28564
Conc: 27.98 ng/ml




#22 AR-1248-2

R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 86446
Conc: 26.05 ng/ml



#22 AR-1248-2


R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 59076
Conc: 40.78 ng/ml



#23 AR-1248-3


R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 102335
Conc: 29.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 52419
Conc: 35.44 ng/ml




#24 AR-1248-4

R.T.: 6.585 min
Delta R.T.: 0.000 min
Response: 105719
Conc: 31.87 ng/ml



#24 AR-1248-4


R.T.: 5.130 min
Delta R.T.: 0.000 min
Response: 60738
Conc: 34.79 ng/ml



#25 AR-1248-5


R.T.: 6.624 min
Delta R.T.: 0.000 min
Response: 87570
Conc: 27.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 55647
Conc: 34.00 ng/ml




#26 AR-1254-1

R.T.: 6.557 min
Delta R.T.: -0.002 min
Response: 91548
Conc: 26.78 ng/ml



#26 AR-1254-1


R.T.: 5.483 min
Delta R.T.: -0.001 min
Response: 83324
Conc: 32.68 ng/ml



#27 AR-1254-2


R.T.: 6.783 min
Delta R.T.: 0.003 min
Response: 132104
Conc: 25.99 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.632 min
Delta R.T.: 0.000 min
Response: 24702
Conc: 10.58 ng/ml




#28 AR-1254-3

R.T.: 7.151 min
Delta R.T.: 0.000 min
Response: 113577
Conc: 22.55 ng/ml



#28 AR-1254-3


R.T.: 6.036 min
Delta R.T.: 0.000 min
Response: 36759
Conc: 10.17 ng/ml



#29 AR-1254-4


R.T.: 7.442 min
Delta R.T.: 0.000 min
Response: 151083
Conc: 41.31 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.265 min
Delta R.T.: -0.001 min
Response: 45903
Conc: 20.59 ng/ml




#30 AR-1254-5

R.T.: 7.866 min
Delta R.T.: 0.000 min
Response: 99023
Conc: 25.30 ng/ml



#30 AR-1254-5


R.T.: 6.686 min
Delta R.T.: 0.000 min
Response: 8714
Conc: 2.59 ng/ml



#31 AR-1260-1


R.T.: 7.321 min
Delta R.T.: 0.002 min
Response: 22659
Conc: 5.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.179 min
Delta R.T.: 0.010 min
Response: 36247
Conc: 13.81 ng/ml




#32 AR-1260-2

R.T.: 7.596 min
Delta R.T.: 0.017 min
Response: 83374
Conc: 17.39 ng/ml



#32 AR-1260-2


R.T.: 6.358 min
Delta R.T.: 0.000 min
Response: 6230
Conc: 1.99 ng/ml



#33 AR-1260-3


R.T.: 7.941 min
Delta R.T.: -0.001 min
Response: 10863
Conc: 3.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.497 min
Delta R.T.: -0.014 min
Response: 18178
Conc: 5.87 ng/ml




#34 AR-1260-4

R.T.: 8.176 min
Delta R.T.: 0.002 min
Response: 11495
Conc: 3.01 ng/ml



#34 AR-1260-4


R.T.: 7.021 min
Delta R.T.: 0.036 min
Response: 7443
Conc: 3.38 ng/ml



#35 AR-1260-5


R.T.: 8.504 min
Delta R.T.: -0.002 min
Response: 15049
Conc: 2.19 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.228 min
Delta R.T.: -0.001 min
Response: 10011
Conc: 2.01 ng/ml




#36 AR-1262-1

R.T.: 7.941 min
Delta R.T.: 0.000 min
Response: 10863
Conc: 2.38 ng/ml



#36 AR-1262-1


R.T.: 6.686 min
Delta R.T.: 0.002 min
Response: 8714
Conc: 5.07 ng/ml



#37 AR-1262-2


R.T.: 8.504 min
Delta R.T.: 0.000 min
Response: 15049
Conc: 1.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.021 min
Delta R.T.: 0.038 min
Response: 7443
Conc: 2.58 ng/ml




#38 AR-1262-3

R.T.: 8.841 min
Delta R.T.: 0.009 min
Response: 5621
Conc: 1.05 ng/ml



#38 AR-1262-3


R.T.: 7.542 min
Delta R.T.: 0.031 min
Response: 9206
Conc: 4.07 ng/ml



#39 AR-1262-4


R.T.: 8.941 min
Delta R.T.: 0.014 min
Response: 6253
Conc: 2.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.574 min
Delta R.T.: -0.002 min
Response: 1073
Conc: 0.26 ng/ml




#40 AR-1262-5

R.T.: 9.613 min
Delta R.T.: 0.000 min
Response: 2116
Conc: 0.76 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.072 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.841 min
Delta R.T.: 0.013 min
Response: 5621
Conc: 0.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.542 min
Delta R.T.: 0.029 min
Response: 9206
Conc: 1.40 ng/ml




#42 AR-1268-2

R.T.: 8.941 min
Delta R.T.: 0.012 min
Response: 6253
Conc: 0.74 ng/ml



#42 AR-1268-2


R.T.: 7.574 min
Delta R.T.: -0.004 min
Response: 1073
Conc: 0.18 ng/ml



#43 AR-1268-3


R.T.: 9.168 min
Delta R.T.: 0.000 min
Response: 6125
Conc: 0.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.785 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.613 min
Delta R.T.: -0.002 min
Response: 2116
Conc: 0.69 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.073 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.059 min
Delta R.T.: 0.005 min
Response: 4533
Conc: 0.19 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.359 min
Delta R.T.: -0.001 min
Response: 6340
Conc: 0.44 ng/ml