

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0030922\  
 Data File : P0085244.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 10 Mar 2022 7:47  
 Operator : YP\AJ  
 Sample : N1645-09  
 Misc : AR1242 MDL 25 PPB  
 ALS Vial : 49 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 MDL-WATER-03-QT1-2022

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 10 08:06:30 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0030822.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Mar 09 04:29:31 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2 | ng/ml  | ng/ml    |
|-----------------------------|--------|-------|---------|--------|--------|----------|
| -----                       |        |       |         |        |        |          |
| System Monitoring Compounds |        |       |         |        |        |          |
| 1) SA Tetrachlo...          | 4.497  | 3.589 | 2882970 | 908752 | 16.790 | 19.915   |
| 2) SA Decachlor...          | 10.407 | 8.607 | 1273412 | 829361 | 19.446 | 21.784   |
| Target Compounds            |        |       |         |        |        |          |
| 16) L4 AR-1242-1            | 5.686  | 4.679 | 73585   | 33973  | 21.595 | 27.136 # |
| 17) L4 AR-1242-2            | 5.709  | 4.697 | 102224  | 52637  | 20.420 | 29.306 # |
| 18) L4 AR-1242-3            | 5.772  | 4.872 | 74854   | 25880  | 24.651 | 26.228   |
| 19) L4 AR-1242-4            | 5.871  | 4.957 | 55015   | 27672  | 23.352 | 27.061   |
| 20) L4 AR-1242-5            | 6.620  | 5.481 | 45168   | 34181  | 23.873 | 26.620   |
| -----                       |        |       |         |        |        |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0030922\  
Data File : P0085244.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 10 Mar 2022 7:47  
Operator : YP\AJ  
Sample : N1645-09  
Misc : AR1242 MDL 25 PPB  
ALS Vial : 49 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
MDL-WATER-03-QT1-2022

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 10 08:06:30 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0030822.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Mar 09 04:29:31 2022  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

