

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031021\
 Data File : P0076034.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Mar 2021 20:09
 Operator : DD\AJ
 Sample : M1606-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 02:29:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 08:07:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.926	3.941	800263	663319	13.426	14.870
2)	SA Decachlor...	10.926	9.210	294322	316247	8.959	8.697
Target Compounds							
8)	L2 AR-1221-1	5.185	0.000	92549	0	153.384	N.D. #
9)	L2 AR-1221-2	5.279	0.000	13219	0	28.823	N.D. #
28)	L6 AR-1254-3	0.000	6.632	0	374989	N.D.	107.779 #
29)	L6 AR-1254-4	8.063	0.000	106403	0	48.672	N.D. #
30)	L6 AR-1254-5	8.501	0.000	82494	0	36.242	N.D. #
31)	L7 AR-1260-1	7.959	0.000	55163	0	25.933	N.D. #
32)	L7 AR-1260-2	0.000	6.991f	0	35885	N.D.	12.967 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031021\
Data File : P0076034.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Mar 2021 20:09
Operator : DD\AJ
Sample : M1606-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 02:29:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 08:07:39 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.926 min
Delta R.T.: -0.002 min
Response: 800263
Conc: 13.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.941 min
Delta R.T.: -0.002 min
Response: 663319
Conc: 14.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.926 min
Delta R.T.: -0.004 min
Response: 294322
Conc: 8.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.210 min
Delta R.T.: -0.013 min
Response: 316247
Conc: 8.70 ng/ml



#8 AR-1221-1


R.T.: 5.185 min
Delta R.T.: 0.005 min
Response: 92549
Conc: 153.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.202 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.279 min
Delta R.T.: 0.000 min
Response: 13219
Conc: 28.82 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.301 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 7.777 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.632 min
Delta R.T.: -0.009 min
Response: 374989
Conc: 107.78 ng/ml




#29 AR-1254-4

R.T.: 8.063 min
Delta R.T.: -0.012 min
Response: 106403
Conc: 48.67 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.880 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.501 min
Delta R.T.: -0.003 min
Response: 82494
Conc: 36.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.308 min
Response: 0
Conc: N.D.




#31 AR-1260-1

R.T.: 7.959 min
Delta R.T.: 0.012 min
Response: 55163
Conc: 25.93 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T. : 6.778 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 8.213 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#32 AR-1260-2

R.T.: 6.991 min
Delta R.T.: 0.016 min
Response: 35885
Conc: 12.97 ng/ml