

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031121\
 Data File : P0076056.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2021 14:09
 Operator : DD\AJ
 Sample : M1458-07
 Misc : AR1242 LOD 25 PPB
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2021

Manual Integrations
 APPROVED

Ankita
 3/12/2021 11:39:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 16:23:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 08:07:39 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.925	3.940	1080303	857521	18.124	19.223
2) SA Decachlor...	10.931	9.212	753366	810645	22.933	22.294
Target Compounds						
16) L4 AR-1242-1	6.248	5.210	13833	32386	9.503	27.229 #
17) L4 AR-1242-2	6.274	5.210	23987	32386	10.753	16.673 #
18) L4 AR-1242-3	6.339	5.399	14481	11255	10.606m	11.498m
19) L4 AR-1242-4	6.444	5.494	11364	10656	10.663m	10.862m
20) L4 AR-1242-5	7.233	6.057	18905	11485	17.640m	9.712m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031121\
Data File : P0076056.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2021 14:09
Operator : DD\AJ
Sample : M1458-07
Misc : AR1242 LOD 25 PPB
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT1-2021

Manual Integrations
APPROVED

Ankita
3/12/2021 11:39:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 16:23:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 08:07:39 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

