

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031125\
 Data File : P0109806.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2025 22:07
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 22:45:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 04:40:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.695	3.693	519.6E6	293.9E6	54.891	56.144
2) SA Decachlor...	8.750	8.702	376.4E6	139.5E6	43.761	43.794
Target Compounds						
16) L4 AR-1242-1	4.788	4.775	142.6E6	75779027	544.712	572.579
17) L4 AR-1242-2	4.808	4.795	195.5E6	105.6E6	549.201	581.759
18) L4 AR-1242-3	4.864	4.970	138.5E6	58517506	550.852	585.910
19) L4 AR-1242-4	4.985	5.054	108.4E6	60038742	551.964	561.538
20) L4 AR-1242-5	5.638	5.576	118.0E6	72386302	550.367	575.841



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031125\
Data File : PO109806.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 22:07
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 22:45:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.695 min
Delta R.T.: -0.003 min
Response: 519551053
Conc: 54.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: -0.001 min
Response: 293868964
Conc: 56.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.750 min
Delta R.T.: -0.007 min
Response: 376424071
Conc: 43.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: -0.005 min
Response: 139450780
Conc: 43.79 ng/ml



#16 AR-1242-1


R.T.: 4.788 min
Delta R.T.: -0.004 min
Response: 142586510
Conc: 544.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.775 min
Delta R.T.: -0.002 min
Response: 75779027
Conc: 572.58 ng/ml




#17 AR-1242-2

R.T.: 4.808 min
Delta R.T.: -0.003 min
Response: 195540993
Conc: 549.20 ng/ml



#17 AR-1242-2


R.T.: 4.795 min
Delta R.T.: -0.002 min
Response: 105598435
Conc: 581.76 ng/ml



#18 AR-1242-3


R.T.: 4.864 min
Delta R.T.: -0.004 min
Response: 138467803
Conc: 550.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.970 min
Delta R.T.: -0.002 min
Response: 58517506
Conc: 585.91 ng/ml




#19 AR-1242-4

R.T.: 4.985 min
Delta R.T.: -0.004 min
Response: 108434446
Conc: 551.96 ng/ml



#19 AR-1242-4


R.T.: 5.054 min
Delta R.T.: -0.003 min
Response: 60038742
Conc: 561.54 ng/ml



#20 AR-1242-5

R.T.: 5.638 min
Delta R.T.: -0.004 min
Response: 117985485
Conc: 550.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.576 min
Delta R.T.: -0.003 min
Response: 72386302
Conc: 575.84 ng/ml