

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031224\
 Data File : P0102416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2024 17:07
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 01:34:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 13 01:32:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.270	3.385	370.7E6	254.4E6	95.846	95.183
2)	SA Decachlor...	9.816	8.254	171.0E6	112.9E6	94.462	94.790
Target Compounds							
16)	L4 AR-1242-1	5.415	4.437	70801651	44864486	919.467	912.298
17)	L4 AR-1242-2	5.437	4.454	103.0E6	61925174	926.194	916.118
18)	L4 AR-1242-3	5.498	4.627	65412696	35875974	922.270	919.966
19)	L4 AR-1242-4	5.596	4.708	54337431	41052096	935.509	909.556
20)	L4 AR-1242-5	6.322	5.220	47769800	39876034	912.267	923.258



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031224\
Data File : PO102416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2024 17:07
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 01:34:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031224.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 13 01:32:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.270 min
Delta R.T.: 0.000 min
Response: 370683681
Conc: 95.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.385 min
Delta R.T.: 0.000 min
Response: 254368601
Conc: 95.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.816 min
Delta R.T.: 0.000 min
Response: 170977720
Conc: 94.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.254 min
Delta R.T.: 0.000 min
Response: 112882992
Conc: 94.79 ng/ml



#16 AR-1242-1


R.T.: 5.415 min
Delta R.T.: 0.000 min
Response: 70801651
Conc: 919.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.437 min
Delta R.T.: 0.000 min
Response: 44864486
Conc: 912.30 ng/ml




#17 AR-1242-2

R.T.: 5.437 min
Delta R.T.: 0.000 min
Response: 102966506
Conc: 926.19 ng/ml



#17 AR-1242-2


R.T.: 4.454 min
Delta R.T.: 0.000 min
Response: 61925174
Conc: 916.12 ng/ml



#18 AR-1242-3


R.T.: 5.498 min
Delta R.T.: 0.000 min
Response: 65412696
Conc: 922.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.627 min
Delta R.T.: 0.000 min
Response: 35875974
Conc: 919.97 ng/ml




#19 AR-1242-4

R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 54337431
Conc: 935.51 ng/ml



#19 AR-1242-4


R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 41052096
Conc: 909.56 ng/ml



#20 AR-1242-5

R.T.: 6.322 min
Delta R.T.: 0.000 min
Response: 47769800
Conc: 912.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.220 min
Delta R.T.: 0.000 min
Response: 39876034
Conc: 923.26 ng/ml