

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031224\
 Data File : P0102423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2024 19:07
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 02:03:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 13 02:00:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.270	3.384	190.5E6	130.8E6	50.000	50.000
2) SA Decachlor...	9.814	8.253	90733757	59989268	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.415	4.436	29091954	19154138	500.000	500.000
22) L5 AR-1248-2	5.684	4.667	50820289	32022404	500.000	500.000
23) L5 AR-1248-3	5.885	4.707	52899889	33393927	500.000	500.000
24) L5 AR-1248-4	6.283	4.875	46198816	37074707	500.000	500.000
25) L5 AR-1248-5	6.322	5.260	44619558	27448345	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031224\
Data File : PO102423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2024 19:07
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 02:03:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031224.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 13 02:00:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.270 min
Delta R.T.: 0.000 min
Response: 190451959
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.384 min
Delta R.T.: 0.000 min
Response: 130829252
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.814 min
Delta R.T.: 0.000 min
Response: 90733757
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.253 min
Delta R.T.: 0.000 min
Response: 59989268
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.415 min
Delta R.T.: 0.000 min
Response: 29091954
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.436 min
Delta R.T.: 0.000 min
Response: 19154138
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 50820289
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 4.667 min
Delta R.T.: 0.000 min
Response: 32022404
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 5.885 min
Delta R.T.: 0.000 min
Response: 52899889
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 4.707 min
Delta R.T.: 0.000 min
Response: 33393927
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.283 min
Delta R.T.: 0.000 min
Response: 46198816
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 37074707
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.322 min
Delta R.T.: 0.000 min
Response: 44619558
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.260 min
Delta R.T.: 0.000 min
Response: 27448345
Conc: 500.00 ng/ml