

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031323\
 Data File : P0093071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2023 19:44
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 04:32:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 04:26:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.429	3.634	237.9E6	85842820	74.385	74.547
2) SA Decachlor...	10.266	8.669	158.7E6	71472401	74.315	74.177
Target Compounds						
26) L6 AR-1254-1	6.469	5.524	82750706	44996626	738.748	740.302
27) L6 AR-1254-2	6.691	5.673	122.7E6	40148877	741.602	741.909
28) L6 AR-1254-3	7.061	6.078	122.1E6	62414514	741.100	740.841
29) L6 AR-1254-4	7.348	6.308	78578343	33597139	740.465	741.695
30) L6 AR-1254-5	7.772	6.728	94060883	54148689	703.942	743.470



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031323\
Data File : PO093071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 19:44
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 04:32:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 04:26:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 237939086
Conc: 74.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: 0.000 min
Response: 85842820
Conc: 74.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.266 min
Delta R.T.: 0.000 min
Response: 158690309
Conc: 74.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.669 min
Delta R.T.: 0.000 min
Response: 71472401
Conc: 74.18 ng/ml



#26 AR-1254-1


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 82750706
Conc: 738.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 44996626
Conc: 740.30 ng/ml




#27 AR-1254-2

R.T.: 6.691 min
Delta R.T.: 0.000 min
Response: 122680293
Conc: 741.60 ng/ml



#27 AR-1254-2


R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 40148877
Conc: 741.91 ng/ml



#28 AR-1254-3


R.T.: 7.061 min
Delta R.T.: 0.000 min
Response: 122071025
Conc: 741.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 62414514
Conc: 740.84 ng/ml




#29 AR-1254-4

R.T.: 7.348 min
Delta R.T.: 0.000 min
Response: 78578343
Conc: 740.47 ng/ml



#29 AR-1254-4


R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 33597139
Conc: 741.69 ng/ml



#30 AR-1254-5

R.T.: 7.772 min
Delta R.T.: 0.000 min
Response: 94060883
Conc: 703.94 ng/ml

Instrument :
ECD_O
ClientSampleId :



#30 AR-1254-5

R.T.: 6.728 min
Delta R.T.: 0.000 min
Response: 54148689
Conc: 743.47 ng/ml