

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031323\
 Data File : P0093084.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2023 23:27
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031323AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/14/2023
 Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 04:40:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 04:39:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.430	3.635	163.2E6	59342348	52.100	52.299
2) SA Decachlor...	10.266	8.668	113.2E6	51764896	51.940	51.399
Target Compounds						
26) L6 AR-1254-1	6.470	5.525	59265572	32225346	507.149	513.439
27) L6 AR-1254-2	6.691	5.673	86649975	28889556	507.117	512.665
28) L6 AR-1254-3	7.060	6.078	85495504	44537693	514.296	517.315
29) L6 AR-1254-4	7.348	6.308	54611633	23529530	512.311	514.522
30) L6 AR-1254-5	7.771	6.728	65333706	38521494	482.493m	521.604

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031323\
Data File : PO093084.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 23:27
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
ICVPO031323AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 04:40:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 04:39:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 163244878
Conc: 52.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031323AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: 0.000 min
Response: 59342348
Conc: 52.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.266 min
Delta R.T.: 0.000 min
Response: 113235292
Conc: 51.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.668 min
Delta R.T.: 0.000 min
Response: 51764896
Conc: 51.40 ng/ml




#26 AR-1254-1

R.T.: 6.470 min
Delta R.T.: 0.001 min
Response: 59265572
Conc: 507.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031323AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023




#26 AR-1254-1

R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 32225346
Conc: 513.44 ng/ml




#27 AR-1254-2

R.T.: 6.691 min
Delta R.T.: 0.000 min
Response: 86649975
Conc: 507.12 ng/ml



#27 AR-1254-2

R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 28889556
Conc: 512.66 ng/ml




#28 AR-1254-3

R.T.: 7.060 min
Delta R.T.: 0.000 min
Response: 85495504
Conc: 514.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031323AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023




#28 AR-1254-3

R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 44537693
Conc: 517.31 ng/ml




#29 AR-1254-4

R.T.: 7.348 min
Delta R.T.: 0.000 min
Response: 54611633
Conc: 512.31 ng/ml



#29 AR-1254-4

R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 23529530
Conc: 514.52 ng/ml




#30 AR-1254-5

R.T.: 7.771 min
Delta R.T.: 0.000 min
Response: 65333706
Conc: 482.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031323AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023



#30 AR-1254-5

R.T.: 6.728 min
Delta R.T.: 0.001 min
Response: 38521494
Conc: 521.60 ng/ml