

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031419\
Data File : P0054391.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2019 15:25
Operator : SM/SJ
Sample : K1861-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BF-10-031119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 16:32:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031419.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 14 14:30:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.013	3.105	86533075	4613842	17.571	18.158
2) SA Decachlor...	9.319	7.666	19534710	4529257	14.181	15.315

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031419\
Data File : P0054391.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2019 15:25
Operator : SM/SJ
Sample : K1861-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BF-10-031119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 16:32:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031419.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 14 14:30:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.013 min
Delta R.T.: -0.001 min
Response: 86533075
Conc: 17.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
BF-10-031119




#1 Tetrachloro-m-xylene

R.T.: 3.105 min
Delta R.T.: 0.000 min
Response: 4613842
Conc: 18.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.319 min
Delta R.T.: 0.000 min
Response: 19534710
Conc: 14.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.666 min
Delta R.T.: 0.000 min
Response: 4529257
Conc: 15.31 ng/ml