

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031419\
 Data File : P0054394.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Mar 2019 16:22
 Operator : SM/SJ
 Sample : K1693-05
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OR-02-031219-A

Manual Integrations
 APPROVED

Sohil
 3/15/2019 3:18:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 00:40:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 14 14:30:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.021	3.106	110.4E6	5919904	22.427m	23.298
2) SA Decachlor...	9.331	7.669	18710269	3350201	13.582	11.328

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031419\
Data File : P0054394.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2019 16:22
Operator : SM/SJ
Sample : K1693-05
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
OR-02-031219-A


Manual Integrations
APPROVED

Sohil
3/15/2019 3:18:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 00:40:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031419.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 14 14:30:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.021 min
Delta R.T.: 0.007 min
Response: 110447503
Conc: 22.43 ng/ml m

Instrument :
ECD_O
ClientSampleId :
OR-02-031219-A


Manual Integrations
APPROVED

Sohil
3/15/2019 3:18:09 PM




#1 Tetrachloro-m-xylene

R.T.: 3.106 min
Delta R.T.: 0.002 min
Response: 5919904
Conc: 23.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.331 min
Delta R.T.: 0.011 min
Response: 18710269
Conc: 13.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.669 min
Delta R.T.: 0.003 min
Response: 3350201
Conc: 11.33 ng/ml