

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031521\
 Data File : P0076160.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Mar 2021 19:55
 Operator : DD\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 05:37:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 16 05:36:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.924	3.938	4181434	2886779	70.686	68.646
2)	SA Decachlor...	10.926	9.205	2827177	2651755	72.476	71.296
Target Compounds							
21)	L5 AR-1248-1	6.248	5.187	662333	484562	683.594	685.283
22)	L5 AR-1248-2	6.542	5.447	1051229	850895	681.387	670.708
23)	L5 AR-1248-3	6.760	5.491	1243137	855558	692.811	675.031
24)	L5 AR-1248-4	7.189	5.674	1294502	996791	704.300	677.401
25)	L5 AR-1248-5	7.230	6.098	1361960	949448	718.528	576.620



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031521\
Data File : P0076160.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Mar 2021 19:55
Operator : DD\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 05:37:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 16 05:36:10 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.924 min
Delta R.T.: 0.002 min
Response: 4181434
Conc: 70.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.938 min
Delta R.T.: 0.002 min
Response: 2886779
Conc: 68.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.926 min
Delta R.T.: 0.005 min
Response: 2827177
Conc: 72.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.205 min
Delta R.T.: 0.002 min
Response: 2651755
Conc: 71.30 ng/ml



#21 AR-1248-1


R.T.: 6.248 min
Delta R.T.: 0.002 min
Response: 662333
Conc: 683.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 5.187 min
Delta R.T.: 0.001 min
Response: 484562
Conc: 685.28 ng/ml




#22 AR-1248-2

R.T.: 6.542 min
Delta R.T.: 0.002 min
Response: 1051229
Conc: 681.39 ng/ml



#22 AR-1248-2


R.T.: 5.447 min
Delta R.T.: 0.002 min
Response: 850895
Conc: 670.71 ng/ml



#23 AR-1248-3


R.T.: 6.760 min
Delta R.T.: 0.002 min
Response: 1243137
Conc: 692.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.491 min
Delta R.T.: 0.001 min
Response: 855558
Conc: 675.03 ng/ml




#24 AR-1248-4

R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 1294502
Conc: 704.30 ng/ml



#24 AR-1248-4


R.T.: 5.674 min
Delta R.T.: 0.001 min
Response: 996791
Conc: 677.40 ng/ml



#25 AR-1248-5

R.T.: 7.230 min
Delta R.T.: 0.002 min
Response: 1361960
Conc: 718.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 6.098 min
Delta R.T.: 0.002 min
Response: 949448
Conc: 576.62 ng/ml