

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031522\
 Data File : P0085359.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Mar 2022 1:00
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 10:42:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 04:43:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.588	981018	222071	4.798	5.183
2)	SA Decachlor...	10.392	8.603	1015136	370999	8.445	9.663
Target Compounds							
3)	L1 AR-1016-1	5.677	0.000	1383	0	0.220	N.D. #
4)	L1 AR-1016-2	5.710	0.000	6736	0	0.735	N.D. #
5)	L1 AR-1016-3	5.762	0.000	8782	0	1.613	N.D. #
6)	L1 AR-1016-4	5.869	0.000	3646	0	0.826	N.D. #
7)	L1 AR-1016-5	6.165	0.000	3148	0	0.741	N.D. #
8)	L2 AR-1221-1	4.706	0.000	1862	0	0.844	N.D. #
9)	L2 AR-1221-2	4.793	3.886	2946	3925	1.808	9.348 #
10)	L2 AR-1221-3	4.870	0.000	870	0	0.169	N.D. #
11)	L3 AR-1232-1	4.870	0.000	870	0	0.185	N.D. #
12)	L3 AR-1232-2	5.415	0.000	8666	0	3.760	N.D. #
13)	L3 AR-1232-3	5.710	0.000	6736	0	1.596	N.D. #
14)	L3 AR-1232-4	5.869	0.000	3646	0	1.816	N.D. #
15)	L3 AR-1232-5	5.952	0.000	2117	0	1.438	N.D. #
16)	L4 AR-1242-1	5.677	0.000	1383	0	0.272	N.D. #
17)	L4 AR-1242-2	5.710	0.000	6736	0	0.907	N.D. #
18)	L4 AR-1242-3	5.762	0.000	8782	0	1.998	N.D. #
19)	L4 AR-1242-4	5.869	0.000	3646	0	1.005	N.D. #
20)	L4 AR-1242-5	6.615	0.000	19383	0	5.455	N.D. #
21)	L5 AR-1248-1	5.677	0.000	1383	0	0.371	N.D. #
22)	L5 AR-1248-2	5.952	0.000	2117	0	0.399	N.D. #
23)	L5 AR-1248-3	6.165	0.000	3148	0	0.535	N.D. #
24)	L5 AR-1248-4	6.564	0.000	4507	0	0.699	N.D. #
25)	L5 AR-1248-5	6.615	0.000	19383	0	3.131	N.D. #
26)	L6 AR-1254-1	6.556	0.000	30258	0	4.566	N.D. #
27)	L6 AR-1254-2	6.769	0.000	14026	0	1.402	N.D. #
28)	L6 AR-1254-3	7.140	0.000	8936	0	0.868	N.D. #
29)	L6 AR-1254-4	7.422	0.000	3881	0	0.553	N.D. #
30)	L6 AR-1254-5	7.854	6.719f	3551	74082	0.464	22.775 #
31)	L7 AR-1260-1	7.329	0.000	48932	0	7.092	N.D. #
32)	L7 AR-1260-2	7.582	0.000	14400	0	1.841	N.D. #
33)	L7 AR-1260-3	7.931	6.509	197759	7111	33.669	2.705 #
34)	L7 AR-1260-4	8.181	6.976	261198	99938	37.816	44.619
35)	L7 AR-1260-5	8.494	7.222	131522	40069	9.362	7.954
36)	L8 AR-1262-1	7.931	6.719f	197759	74082	21.896	43.147 #
37)	L8 AR-1262-2	8.494	6.976	131522	99938	8.254	34.897 #
38)	L8 AR-1262-3	8.815	7.507	892832	367712	86.380	167.904 #
39)	L8 AR-1262-4	8.912	7.572	802290	319073	170.815	79.956 #
40)	L8 AR-1262-5	9.597	8.067	282923	110976	53.201	60.528
41)	L9 AR-1268-1	8.815	7.507	892832	367712	46.262	55.367

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031522\
Data File : P0085359.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2022 1:00
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 10:42:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.912	7.572	802290	319073	46.173	54.136
43) L9	AR-1268-3	9.152	7.779	670688	269680	45.441	54.279
44) L9	AR-1268-4	9.597	8.067	282923	110976	46.370	52.294
45) L9	AR-1268-5	10.035	8.354	2169379	726907	44.938	51.905



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031522\
Data File : P0085359.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2022 1:00
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 10:42:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: 0.000 min
Response: 981018
Conc: 4.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.001 min
Response: 222071
Conc: 5.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.392 min
Delta R.T.: -0.004 min
Response: 1015136
Conc: 8.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.603 min
Delta R.T.: -0.003 min
Response: 370999
Conc: 9.66 ng/ml



#3 AR-1016-1


R.T.: 5.677 min
Delta R.T.: -0.004 min
Response: 1383
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.679 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.710 min
Delta R.T.: 0.007 min
Response: 6736
Conc: 0.74 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.697 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.762 min
Delta R.T.: -0.005 min
Response: 8782
Conc: 1.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.872 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.869 min
Delta R.T.: 0.003 min
Response: 3646
Conc: 0.83 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.916 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 3148
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.129 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.706 min
Delta R.T.: 0.003 min
Response: 1862
Conc: 0.84 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.805 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.793 min
Delta R.T.: 0.002 min
Response: 2946
Conc: 1.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#9 AR-1221-2

R.T.: 3.886 min
Delta R.T.: -0.004 min
Response: 3925
Conc: 9.35 ng/ml




#10 AR-1221-3

R.T.: 4.870 min
Delta R.T.: 0.001 min
Response: 870
Conc: 0.17 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.966 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.870 min
Delta R.T.: 0.001 min
Response: 870
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.966 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.415 min
Delta R.T.: 0.007 min
Response: 8666
Conc: 3.76 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.697 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.710 min
Delta R.T.: 0.006 min
Response: 6736
Conc: 1.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.872 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.869 min
Delta R.T.: 0.002 min
Response: 3646
Conc: 1.82 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.957 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.952 min
Delta R.T.: -0.007 min
Response: 2117
Conc: 1.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.128 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 1383
Conc: 0.27 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.678 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.710 min
Delta R.T.: 0.007 min
Response: 6736
Conc: 0.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.696 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.762 min
Delta R.T.: -0.004 min
Response: 8782
Conc: 2.00 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.872 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.869 min
Delta R.T.: 0.002 min
Response: 3646
Conc: 1.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.957 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.615 min
Delta R.T.: 0.001 min
Response: 19383
Conc: 5.45 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T. : 5.480 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.677 min
Delta R.T.: -0.002 min
Response: 1383
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.678 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.952 min
Delta R.T.: -0.005 min
Response: 2117
Conc: 0.40 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.915 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 3148
Conc: 0.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.956 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.564 min
Delta R.T.: -0.009 min
Response: 4507
Conc: 0.70 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.127 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.615 min
Delta R.T.: 0.002 min
Response: 19383
Conc: 3.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.521 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.556 min
Delta R.T.: 0.009 min
Response: 30258
Conc: 4.57 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.480 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.769 min
Delta R.T.: 0.000 min
Response: 14026
Conc: 1.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.628 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.140 min
Delta R.T.: 0.001 min
Response: 8936
Conc: 0.87 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.032 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.422 min
Delta R.T.: -0.005 min
Response: 3881
Conc: 0.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.262 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.854 min
Delta R.T.: 0.002 min
Response: 3551
Conc: 0.46 ng/ml



#30 AR-1254-5


R.T.: 6.719 min
Delta R.T.: 0.039 min
Response: 74082
Conc: 22.77 ng/ml



#31 AR-1260-1


R.T.: 7.329 min
Delta R.T.: 0.022 min
Response: 48932
Conc: 7.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.165 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.582 min
Delta R.T.: 0.016 min
Response: 14400
Conc: 1.84 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.355 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.931 min
Delta R.T.: 0.000 min
Response: 197759
Conc: 33.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#33 AR-1260-3

R.T.: 6.509 min
Delta R.T.: 0.002 min
Response: 7111
Conc: 2.70 ng/ml




#34 AR-1260-4

R.T.: 8.181 min
Delta R.T.: 0.021 min
Response: 261198
Conc: 37.82 ng/ml



#34 AR-1260-4


R.T.: 6.976 min
Delta R.T.: -0.005 min
Response: 99938
Conc: 44.62 ng/ml



#35 AR-1260-5


R.T.: 8.494 min
Delta R.T.: 0.002 min
Response: 131522
Conc: 9.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#35 AR-1260-5

R.T.: 7.222 min
Delta R.T.: -0.003 min
Response: 40069
Conc: 7.95 ng/ml




#36 AR-1262-1

R.T.: 7.931 min
Delta R.T.: 0.001 min
Response: 197759
Conc: 21.90 ng/ml



#36 AR-1262-1


R.T.: 6.719 min
Delta R.T.: 0.040 min
Response: 74082
Conc: 43.15 ng/ml



#37 AR-1262-2


R.T.: 8.494 min
Delta R.T.: 0.004 min
Response: 131522
Conc: 8.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#37 AR-1262-2

R.T.: 6.976 min
Delta R.T.: -0.003 min
Response: 99938
Conc: 34.90 ng/ml




#38 AR-1262-3

R.T.: 8.815 min
Delta R.T.: -0.003 min
Response: 892832
Conc: 86.38 ng/ml



#38 AR-1262-3


R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 367712
Conc: 167.90 ng/ml



#39 AR-1262-4


R.T.: 8.912 min
Delta R.T.: 0.002 min
Response: 802290
Conc: 170.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#39 AR-1262-4

R.T.: 7.572 min
Delta R.T.: 0.000 min
Response: 319073
Conc: 79.96 ng/ml




#40 AR-1262-5

R.T.: 9.597 min
Delta R.T.: 0.000 min
Response: 282923
Conc: 53.20 ng/ml



#40 AR-1262-5


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 110976
Conc: 60.53 ng/ml



#41 AR-1268-1


R.T.: 8.815 min
Delta R.T.: 0.000 min
Response: 892832
Conc: 46.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 367712
Conc: 55.37 ng/ml




#42 AR-1268-2

R.T.: 8.912 min
Delta R.T.: 0.001 min
Response: 802290
Conc: 46.17 ng/ml



#42 AR-1268-2


R.T.: 7.572 min
Delta R.T.: 0.000 min
Response: 319073
Conc: 54.14 ng/ml



#43 AR-1268-3


R.T.: 9.152 min
Delta R.T.: 0.001 min
Response: 670688
Conc: 45.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.779 min
Delta R.T.: 0.000 min
Response: 269680
Conc: 54.28 ng/ml




#44 AR-1268-4

R.T.: 9.597 min
Delta R.T.: 0.000 min
Response: 282923
Conc: 46.37 ng/ml



#44 AR-1268-4


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 110976
Conc: 52.29 ng/ml



#45 AR-1268-5

R.T.: 10.035 min
Delta R.T.: 0.001 min
Response: 2169379
Conc: 44.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.354 min
Delta R.T.: 0.000 min
Response: 726907
Conc: 51.91 ng/ml