

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031722.M
 Title : GC EXTRACTABLES
 Last Update : Thu Mar 17 03:20:00 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP044831.D 1000 =PP044829.D 750 =PP044830.D
 250 =PP044832.D 50 =PP044833.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.040	1.981	2.019	2.050	1.689	1.956 E6	7.74
2) SA	Decachlorobiphenyl	1.391	1.310	1.347	1.431	1.316	1.359 E6	3.80
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	6.692	6.295	6.463	6.886	6.864	6.640 E4	3.87
27) L6	AR-1254-2	1.001	0.949	0.977	1.052	1.016	0.999 E5	3.89
28) L6	AR-1254-3	1.031	0.987	1.006	1.069	1.021	1.023 E5	3.02
29) L6	AR-1254-4	7.294	6.982	7.155	7.588	7.349	7.273 E4	3.11
30) L6	AR-1254-5	7.852	7.581	7.742	8.002	7.494	7.734 E4	2.64
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PP044831.D 1000 =PP044829.D 750 =PP044830.D
 250 =PP044832.D 50 =PP044833.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.508	1.471	1.508	1.530	1.391	1.482 E6	3.71
2) SA	Decachlorobiphenyl	1.426	1.328	1.372	1.518	1.501	1.429 E6	5.72
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031722.M
 Title : GC EXTRACTABLES
 Last Update : Thu Mar 17 03:20:00 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP044831.D 1000 =PP044829.D 750 =PP044830.D
 250 =PP044832.D 50 =PP044833.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	8.495	7.969	8.226	9.065	9.483	8.648 E4	7.16
27) L6	AR-1254-2	7.329	6.816	7.074	7.878	8.109	7.441 E4	7.29
28) L6	AR-1254-3	1.146	1.089	1.121	1.209	1.180	1.149 E5	4.11
29) L6	AR-1254-4	7.467	7.020	7.243	7.897	7.650	7.456 E4	4.59
30) L6	AR-1254-5	1.041	0.978	1.010	1.096	1.024	1.030 E5	4.24
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range