

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031722.M
 Title : GC EXTRACTABLES
 Last Update : Thu Mar 17 03:59:19 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP044837.D 1000 =PP044835.D 750 =PP044836.D
 250 =PP044838.D 50 =PP044839.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.039	1.963	1.998	2.047	1.724	1.954 E6	6.81
2) SA	Decachlorobiphenyl	2.364	2.218	2.270	2.440	2.193	2.297 E6	4.51
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.982	1.883	1.918	2.045	1.796	1.925 E5	4.93
42) L9	AR-1268-2	1.792	1.705	1.752	1.843	1.638	1.746 E5	4.53
43) L9	AR-1268-3	1.579	1.494	1.526	1.605	1.432	1.527 E5	4.50
44) L9	AR-1268-4	6.488	6.263	6.330	6.555	5.658	6.259 E4	5.68
45) L9	AR-1268-5	4.952	4.783	4.834	5.083	4.440	4.819 E5	5.00

Signal #2 Calibration Files

500 =PP044837.D 1000 =PP044835.D 750 =PP044836.D
 250 =PP044838.D 50 =PP044839.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.512	1.472	1.487	1.527	1.328	1.465 E6	5.43
2) SA	Decachlorobiphenyl	2.425	2.278	2.337	2.548	2.512	2.420 E6	4.70
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031722.M
 Title : GC EXTRACTABLES
 Last Update : Thu Mar 17 03:59:19 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP044837.D 1000 =PP044835.D 750 =PP044836.D
 250 =PP044838.D 50 =PP044839.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.154	2.064	2.087	2.272	2.213	2.158 E5	4.00
42) L9	AR-1268-2	1.967	1.874	1.913	2.037	1.986	1.956 E5	3.24
43) L9	AR-1268-3	1.657	1.571	1.606	1.724	1.677	1.647 E5	3.66
44) L9	AR-1268-4	7.654	7.143	7.338	8.103	8.057	7.659 E4	5.56
45) L9	AR-1268-5	5.240	5.034	5.110	5.364	5.017	5.153 E5	2.85

(#) = Out of Range