

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031622\
 Data File : P0085403.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Mar 2022 17:19
 Operator : YP\AJ
 Sample : PB143386BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB143386BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 00:55:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 04:43:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.590	3448178	792083	16.863	18.488
2) SA Decachlor...	10.394	8.605	2030252	662154	16.890	17.247
Target Compounds						
3) L1 AR-1016-1	5.680	4.678	2541995	645411	403.931	404.424
4) L1 AR-1016-2	5.703	4.697	3669808	922551	400.635	398.722
5) L1 AR-1016-3	5.766	4.872	2227039	519892	409.143	415.992
6) L1 AR-1016-4	5.866	4.915	1835184	417271	416.043	382.244
7) L1 AR-1016-5	6.165	5.128	1720905	527867	405.284	390.263
31) L7 AR-1260-1	7.307	6.164	2814471	997394	407.895	420.695
32) L7 AR-1260-2	7.565	6.353	3149289	1173727	402.587	420.452
33) L7 AR-1260-3	7.930	6.506	1916673	1088732	326.319	414.094 #
34) L7 AR-1260-4	8.160	6.980	2395231	781249	346.782	348.801
35) L7 AR-1260-5	8.492	7.224	4752201	1784487	338.256	354.218



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031622\
Data File : P0085403.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2022 17:19
Operator : YP\AJ
Sample : PB143386BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB143386BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 00:55:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 3448178
Conc: 16.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: 0.000 min
Response: 792083
Conc: 18.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.394 min
Delta R.T.: -0.002 min
Response: 2030252
Conc: 16.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.605 min
Delta R.T.: 0.000 min
Response: 662154
Conc: 17.25 ng/ml



#3 AR-1016-1


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 2541995
Conc: 403.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#3 AR-1016-1

R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 645411
Conc: 404.42 ng/ml




#4 AR-1016-2

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 3669808
Conc: 400.63 ng/ml



#4 AR-1016-2


R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 922551
Conc: 398.72 ng/ml



#5 AR-1016-3


R.T.: 5.766 min
Delta R.T.: 0.000 min
Response: 2227039
Conc: 409.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#5 AR-1016-3

R.T.: 4.872 min
Delta R.T.: 0.000 min
Response: 519892
Conc: 415.99 ng/ml




#6 AR-1016-4

R.T.: 5.866 min
Delta R.T.: 0.000 min
Response: 1835184
Conc: 416.04 ng/ml



#6 AR-1016-4


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 417271
Conc: 382.24 ng/ml



#7 AR-1016-5


R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 1720905
Conc: 405.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#7 AR-1016-5

R.T.: 5.128 min
Delta R.T.: 0.000 min
Response: 527867
Conc: 390.26 ng/ml




#31 AR-1260-1

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 2814471
Conc: 407.89 ng/ml



#31 AR-1260-1


R.T.: 6.164 min
Delta R.T.: 0.000 min
Response: 997394
Conc: 420.70 ng/ml



#32 AR-1260-2


R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 3149289
Conc: 402.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#32 AR-1260-2

R.T.: 6.353 min
Delta R.T.: -0.001 min
Response: 1173727
Conc: 420.45 ng/ml




#33 AR-1260-3

R.T.: 7.930 min
Delta R.T.: 0.000 min
Response: 1916673
Conc: 326.32 ng/ml



#33 AR-1260-3


R.T.: 6.506 min
Delta R.T.: 0.000 min
Response: 1088732
Conc: 414.09 ng/ml



#34 AR-1260-4


R.T.: 8.160 min
Delta R.T.: 0.000 min
Response: 2395231
Conc: 346.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143386BS




#34 AR-1260-4

R.T.: 6.980 min
Delta R.T.: 0.000 min
Response: 781249
Conc: 348.80 ng/ml



#35 AR-1260-5

R.T.: 8.492 min
Delta R.T.: 0.000 min
Response: 4752201
Conc: 338.26 ng/ml



#35 AR-1260-5

R.T.: 7.224 min
Delta R.T.: -0.001 min
Response: 1784487
Conc: 354.22 ng/ml