

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031623\
 Data File : P0093217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2023 15:55
 Operator : YP/AJ
 Sample : 01932-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CE-81-031423

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/17/2023
 Supervised By :Ankita Jodhani 03/17/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 21:57:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.633	62715676	22684768	20.027	20.113
2) SA Decachlor...	10.261	8.665	37654170	17327754	17.398	17.697

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031623\
Data File : PO093217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2023 15:55
Operator : YP/AJ
Sample : 01932-07
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CE-81-031423


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/17/2023
Supervised By :Ankita Jodhani 03/17/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 21:57:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.427 min
Delta R.T.: -0.003 min
Response: 62715676
Conc: 20.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
CE-81-031423


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/17/2023
Supervised By :Ankita Jodhani 03/17/2023




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: -0.002 min
Response: 22684768
Conc: 20.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: -0.002 min
Response: 37654170
Conc: 17.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.665 min
Delta R.T.: -0.003 min
Response: 17327754
Conc: 17.70 ng/ml