

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031623\
 Data File : P0093251.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Mar 2023 03:06
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 04:22:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.633	153.0E6	55987004	48.861	49.640
2) SA Decachlor...	10.260	8.663	108.6E6	49425053	50.183	50.478
Target Compounds						
21) L5 AR-1248-1	5.604	4.718	29526778	11459944	474.982	492.442
22) L5 AR-1248-2	5.880	4.955	47008025	17561928	492.180	484.593
23) L5 AR-1248-3	6.086	4.996	50749254	18653148	487.696	497.065
24) L5 AR-1248-4	6.493	5.168	50031088	21188694	491.314	492.515
25) L5 AR-1248-5	6.532	5.562	51499495	18824093	503.125	511.839



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031623\
Data File : PO093251.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2023 03:06
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 04:22:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.427 min
Delta R.T.: -0.003 min
Response: 153008718
Conc: 48.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: -0.002 min
Response: 55987004
Conc: 49.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: -0.003 min
Response: 108610471
Conc: 50.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.663 min
Delta R.T.: -0.005 min
Response: 49425053
Conc: 50.48 ng/ml



#21 AR-1248-1


R.T.: 5.604 min
Delta R.T.: -0.002 min
Response: 29526778
Conc: 474.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.718 min
Delta R.T.: -0.003 min
Response: 11459944
Conc: 492.44 ng/ml




#22 AR-1248-2

R.T.: 5.880 min
Delta R.T.: -0.003 min
Response: 47008025
Conc: 492.18 ng/ml



#22 AR-1248-2


R.T.: 4.955 min
Delta R.T.: -0.003 min
Response: 17561928
Conc: 484.59 ng/ml



#23 AR-1248-3


R.T.: 6.086 min
Delta R.T.: -0.003 min
Response: 50749254
Conc: 487.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.996 min
Delta R.T.: -0.003 min
Response: 18653148
Conc: 497.06 ng/ml




#24 AR-1248-4

R.T.: 6.493 min
Delta R.T.: -0.003 min
Response: 50031088
Conc: 491.31 ng/ml



#24 AR-1248-4


R.T.: 5.168 min
Delta R.T.: -0.003 min
Response: 21188694
Conc: 492.51 ng/ml



#25 AR-1248-5

R.T.: 6.532 min
Delta R.T.: -0.003 min
Response: 51499495
Conc: 503.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.562 min
Delta R.T.: -0.003 min
Response: 18824093
Conc: 511.84 ng/ml