Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031822\
Data File : PO085512.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2022 22:03
Operator : YP\AJ
Sample : N1958-04
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
C-4-3-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 06:21:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.491	3.589	3840161	927141	18.780	21.640
2)	SA Decachlor...	10.393	8.606	2021768	634920	16.819	16.537
	Target Compounds						
26)	L6 AR-1254-1	6.548	5.481	3954389	1150584	596.685	437.781 #
27)	L6 AR-1254-2	6.768	5.628	4788461	1129601	478.478	484.104
28)	L6 AR-1254-3	7.139	6.032	4957242	1847114	481.681	502.915
29)	L6 AR-1254-4	7.427	6.262	3758797	1124990	535.754	507.877
30)	L6 AR-1254-5	7.851	6.681	6809025	2471879	889.887	759.928

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031822\
Data File : PO085512.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2022 22:03
Operator : YP\AJ
Sample : N1958-04
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
C-4-3-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 06:21:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm