

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031824\
 Data File : P0102564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 19:39
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031824

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 22:41:00 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 18 22:40:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.267	3.383	213.9E6	142.2E6	52.094	51.289
2) SA Decachlor...	9.809	8.249	123.7E6	84988564	50.772	51.091
Target Compounds						
16) L4 AR-1242-1	5.412	4.434	46699708	29309957	518.344	502.174
17) L4 AR-1242-2	5.434	4.452	66519718	41163354	503.967	504.831
18) L4 AR-1242-3	5.495	4.623	43324170	23008747	510.958	510.752
19) L4 AR-1242-4	5.592	4.705	34626985	26627284	516.152	513.738
20) L4 AR-1242-5	6.318	5.217	30404645	28326236	485.611	505.890



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031824\
Data File : PO102564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 19:39
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031824

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 22:41:00 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 18 22:40:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.267 min
Delta R.T.: 0.000 min
Response: 213932324
Conc: 52.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#1 Tetrachloro-m-xylene

R.T.: 3.383 min
Delta R.T.: 0.000 min
Response: 142199626
Conc: 51.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.809 min
Delta R.T.: 0.000 min
Response: 123687599
Conc: 50.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.249 min
Delta R.T.: 0.000 min
Response: 84988564
Conc: 51.09 ng/ml



#16 AR-1242-1


R.T.: 5.412 min
Delta R.T.: 0.000 min
Response: 46699708
Conc: 518.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#16 AR-1242-1

R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 29309957
Conc: 502.17 ng/ml




#17 AR-1242-2

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 66519718
Conc: 503.97 ng/ml



#17 AR-1242-2


R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 41163354
Conc: 504.83 ng/ml



#18 AR-1242-3


R.T.: 5.495 min
Delta R.T.: 0.000 min
Response: 43324170
Conc: 510.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#18 AR-1242-3

R.T.: 4.623 min
Delta R.T.: 0.000 min
Response: 23008747
Conc: 510.75 ng/ml




#19 AR-1242-4

R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 34626985
Conc: 516.15 ng/ml



#19 AR-1242-4


R.T.: 4.705 min
Delta R.T.: 0.000 min
Response: 26627284
Conc: 513.74 ng/ml



#20 AR-1242-5

R.T.: 6.318 min
Delta R.T.: 0.000 min
Response: 30404645
Conc: 485.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824



#20 AR-1242-5

R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 28326236
Conc: 505.89 ng/ml