

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031824\
 Data File : P0102563.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 19:21
 Operator : YP/AJ
 Sample : P0031824ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031824

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 22:44:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 18 22:43:37 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.267	3.382	219.6E6	146.4E6	49.813	50.896
2) SA Decachlor...	9.810	8.249	131.9E6	90347825	47.499	49.801
Target Compounds						
3) L1 AR-1016-1	5.412	4.434	64187002	39325918	486.735	483.642
4) L1 AR-1016-2	5.433	4.452	93886442	56881608	482.565	491.172
5) L1 AR-1016-3	5.493	4.623	52246714	31067518	490.445	493.393
6) L1 AR-1016-4	5.591	4.664	49584373	29322130	499.910	501.866
7) L1 AR-1016-5	5.881	4.872	48816576	35872297	508.156	501.580
31) L7 AR-1260-1	6.993	5.884	83581496	66577053	495.272	503.326
32) L7 AR-1260-2	7.247	6.071	88888685	66348636	477.182	486.117
33) L7 AR-1260-3	7.603	6.221	59529380	71071149	497.742	487.953
34) L7 AR-1260-4	7.826	6.685	68742223	50283561	482.346	501.159
35) L7 AR-1260-5	8.131	6.927	112.6E6	103.1E6	466.017	485.250



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031824\
Data File : P0102563.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 19:21
Operator : YP/AJ
Sample : P0031824ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031824

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 22:44:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 18 22:43:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.267 min
Delta R.T.: 0.000 min
Response: 219623206
Conc: 49.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#1 Tetrachloro-m-xylene

R.T.: 3.382 min
Delta R.T.: 0.000 min
Response: 146445713
Conc: 50.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.810 min
Delta R.T.: 0.000 min
Response: 131878894
Conc: 47.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.249 min
Delta R.T.: 0.000 min
Response: 90347825
Conc: 49.80 ng/ml



#3 AR-1016-1


R.T.: 5.412 min
Delta R.T.: 0.000 min
Response: 64187002
Conc: 486.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#3 AR-1016-1

R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 39325918
Conc: 483.64 ng/ml




#4 AR-1016-2

R.T.: 5.433 min
Delta R.T.: 0.000 min
Response: 93886442
Conc: 482.56 ng/ml



#4 AR-1016-2


R.T.: 4.452 min
Delta R.T.: 0.000 min
Response: 56881608
Conc: 491.17 ng/ml



#5 AR-1016-3


R.T.: 5.493 min
Delta R.T.: 0.000 min
Response: 52246714
Conc: 490.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#5 AR-1016-3

R.T.: 4.623 min
Delta R.T.: 0.000 min
Response: 31067518
Conc: 493.39 ng/ml




#6 AR-1016-4

R.T.: 5.591 min
Delta R.T.: 0.000 min
Response: 49584373
Conc: 499.91 ng/ml



#6 AR-1016-4


R.T.: 4.664 min
Delta R.T.: 0.000 min
Response: 29322130
Conc: 501.87 ng/ml



#7 AR-1016-5


R.T.: 5.881 min
Delta R.T.: 0.000 min
Response: 48816576
Conc: 508.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#7 AR-1016-5

R.T.: 4.872 min
Delta R.T.: 0.000 min
Response: 35872297
Conc: 501.58 ng/ml




#31 AR-1260-1

R.T.: 6.993 min
Delta R.T.: 0.000 min
Response: 83581496
Conc: 495.27 ng/ml



#31 AR-1260-1


R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 66577053
Conc: 503.33 ng/ml



#32 AR-1260-2


R.T.: 7.247 min
Delta R.T.: 0.000 min
Response: 88888685
Conc: 477.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#32 AR-1260-2

R.T.: 6.071 min
Delta R.T.: 0.000 min
Response: 66348636
Conc: 486.12 ng/ml




#33 AR-1260-3

R.T.: 7.603 min
Delta R.T.: 0.000 min
Response: 59529380
Conc: 497.74 ng/ml



#33 AR-1260-3


R.T.: 6.221 min
Delta R.T.: 0.000 min
Response: 71071149
Conc: 487.95 ng/ml



#34 AR-1260-4


R.T.: 7.826 min
Delta R.T.: 0.000 min
Response: 68742223
Conc: 482.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#34 AR-1260-4

R.T.: 6.685 min
Delta R.T.: 0.000 min
Response: 50283561
Conc: 501.16 ng/ml



#35 AR-1260-5

R.T.: 8.131 min
Delta R.T.: 0.000 min
Response: 112582526
Conc: 466.02 ng/ml



#35 AR-1260-5

R.T.: 6.927 min
Delta R.T.: 0.000 min
Response: 103118176
Conc: 485.25 ng/ml